



RETURN BIDS TO:

RETOURNER LES SOUMISSIONS À:

Bid Receiving Public Works and Government
Services Canada/Réception des soumissions Travaux
publics et Services gouvernementaux Canada
Room 100,
167 Lombard Ave.
Winnipeg
Manitoba
R3B 0T6
Bid Fax: (204) 983-0338

SOLICITATION AMENDMENT MODIFICATION DE L'INVITATION

The referenced document is hereby revised; unless otherwise
indicated, all other terms and conditions of the Solicitation
remain the same.

Ce document est par la présente révisé; sauf indication contraire,
les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

Vendor/Firm Name and Address
Raison sociale et adresse du
fournisseur/de l'entrepreneur

Issuing Office - Bureau de distribution
Public Works and Government Services Canada -
Western Region
Room 100
167 Lombard Ave.
Winnipeg
Manitoba
R3B 0T6

Title - Sujet Neonatal Birthing Beds	
Solicitation No. - N° de l'invitation H3593-172932/A	Amendment No. - N° modif. 007
Client Reference No. - N° de référence du client H3593-172932	Date 2018-12-21
GETS Reference No. - N° de référence de SEAG PW-\$WPG-113-10700	
File No. - N° de dossier WPG-8-41121 (113)	CCC No./N° CCC - FMS No./N° VME
Solicitation Closes - L'invitation prend fin at - à 02:00 PM on - le 2019-01-08	
Time Zone Fuseau horaire Central Standard Time CST	
F.O.B. - F.A.B. Plant-Usine: <input type="checkbox"/> Destination: <input checked="" type="checkbox"/> Other-Autre: <input type="checkbox"/>	
Address Enquiries to: - Adresser toutes questions à: Chubey, Karen	Buyer Id - Id de l'acheteur wpg113
Telephone No. - N° de téléphone (204) 291-5928 ()	FAX No. - N° de FAX (204) 983-7796
Destination - of Goods, Services, and Construction: Destination - des biens, services et construction:	

Instructions: See Herein

Instructions: Voir aux présentes

Delivery Required - Livraison exigée	Delivery Offered - Livraison proposée
Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur	
Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur	
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date

Neonatal Birthing Beds

Amendment No. 007 is raised to answer and amend:

Amendment No. 007 is raised to answer:

QUESTION 1:

Q1: We would greatly appreciate the opportunity to respond with our newest device. This device meets and exceeds most of the technical specs in the document, but not all of them as they are specific.

Is there a possibility to submit our technology by either removing/modifying the mandatory specifications 2.5, 2.10, 2.11, 2.17, 2.18 and 2.19?

RESPONSE 1:

Your request has been given careful consideration however the mandatory specifications remain unchanged.

The specifications included in the compliance matrix are essential as the equipment is used in a high stress environment providing health services in northern Manitoba with space constraints. As described in 1.2 and 1.3 of Annex A the multipurpose neonatal bed is required to fulfill a variety of functions simultaneously in order to reduce the amount of equipment stored in the neonatal unit of the Norway House Hospital.

Further information is below:

2.10 Shuttle system, must be electric with built in battery, powered for intra hospital movements. With minimum 30 minutes continuous power supply:

An all in one unit with built in battery power is required for transporting newborns within the Hospital. In emergency situation loss of power in transport would place patient care at risk. External attachments are cumbersome when transporting in emergency situations and the risk to continuous power supply during transport if the external attachment is misplaced or lost.

2.11 System must be able to provide care to 1 neonatal patient with built in shuttle system: Unsure what is being asked here.

We need a system that can provide care for minimum of one neonatal patient.

2.17 Adaptable Alarm system with hands-free silence: Hands Free alarm silence is proprietary. Alarm management options on our device can decrease if not eliminate the need for alarm silence, decreasing alarm fatigue.

This is an essential specification when we are busy with mom or baby due to limited staff that we have available during emergency situations hands-free is essential.

2.18 Full Display, of temperature, pulse rate, SpO2 and weight: This is only available on the ECG monitor attachment. This spec is from an open-care device only. Not available on any combination bed.

This is an essential specification as we have limited space and an all in one encompass neonatal bed. It is vital with a newborn emergency that we are able to closely monitor with full display, of temperature,

Sollicitation No. – N° de l'invitation
H3593-172932/A
Client Ref. No. - N° de réf. du client
H3593-172932

Amd. No. - N° de la modif.
007
File No. - N° du dossier
7342-H3593-172932

Buyer ID - Id de l'acheteur
wpg113
CCC No./N° CCC - FMS No./N° VME

pulse rate, SpO2 and weight. When your emergency situation with a neonate we have limited work space and having this option close and visible is important.

2.19 Integrated Resuscitation system: Not available on combination bed, open-care devices only. Can side-mount a resuscitation device as needed, which is what most sites do.

This is essential specification to allow staff to respond in an emergency situation providing care to emergent neonate in one system that we provide the necessary emergent life saving measures.

ALL OTHER TERMS AND CONDITIONS REMAIN THE SAME.