



National Defence

Défense nationale

National Defence Headquarters  
Ottawa, Ontario  
K1A 0K2

Quartier général de la Défense nationale  
Ottawa (Ontario)  
K1A 0K2

**SOLICITATION CLOSES  
L'INVITATION PREND FIN**

at - à 2:00 PM  
on - le 18 January 2019

**TIME ZONE - FUSEAU HORAIRE**

**Eastern Standard Time (EST)**

**SOLICITATION AMENDMENT  
MODIFICATION DE L'INVITATION**

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Solicitation remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de l'invitation demeurent les mêmes.

**Comments - Commentaires**

**Issuing Office - Bureau de distribution**

Director of Services Contracting (D Svcs C) 3  
Department of National Defence  
101 Colonel By Drive  
Ottawa ON K1A 0K2

<b>Title - Sujet</b> Calibration, Maintenance and Repair Services for Medical Test Equipment and Simulation Devices					
<b>Solicitation No. - N° de l'invitation</b> W6369-17-A005/B	<b>Amendment No. - N° de la modification</b> 003				
<b>Date of Amendment - Date de modification</b> 11 January 2019					
<b>RETURN BIDS TO: RETOURNER LES SOUMISSIONS À :</b>  By e-mail to: <a href="mailto:Shiroma.Ratnayake@forces.gc.ca">Shiroma.Ratnayake@forces.gc.ca</a>  Director Services Contracting (D Svcs C) 3-2-8 Attention: Shiroma Ratnayake					
<b>Address enquiries to: Adresser toute demande de renseignements à:</b>  Shiroma Ratnayake  <table border="0"> <tr> <td><b>Telephone No. N° de téléphone</b></td> <td><b>E-Mail Address Courriel</b></td> </tr> <tr> <td></td> <td><a href="mailto:Shiroma.Ratnayake@forces.gc.ca">Shiroma.Ratnayake@forces.gc.ca</a></td> </tr> </table>		<b>Telephone No. N° de téléphone</b>	<b>E-Mail Address Courriel</b>		<a href="mailto:Shiroma.Ratnayake@forces.gc.ca">Shiroma.Ratnayake@forces.gc.ca</a>
<b>Telephone No. N° de téléphone</b>	<b>E-Mail Address Courriel</b>				
	<a href="mailto:Shiroma.Ratnayake@forces.gc.ca">Shiroma.Ratnayake@forces.gc.ca</a>				
<b>Destination</b> See herein.					
<b>Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur</b> <b>Telephone No.</b>  <table border="0"> <tr> <td><b>N° de téléphone</b></td> <td><b>E-Mail Address Courriel</b></td> </tr> <tr> <td></td> <td></td> </tr> </table>		<b>N° de téléphone</b>	<b>E-Mail Address Courriel</b>		
<b>N° de téléphone</b>	<b>E-Mail Address Courriel</b>				
<b>Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/de l'entrepreneur (taper ou écrire en caractères d'imprimerie)</b>					
<b>Signature</b>	<b>Date</b>				



**AMENDMENT 003 TO SOLICITATION NUMBER W6369-17-A005/B IS RAISED TO:**

1. Extend the solicitation closing date by four (4) calendar days from 14 January 2019 to 18 January 2019. .
2. Provide answers to all other Bidder Questionnaire.

**ALL OTHER TERMS AND CONDITIONS REMAIN THE SAME.**

<p><b>Question 4:</b></p>	<p>Are any or all the calibration procedures available on the list of equipment? If so can we request the calibration procedures for the following;</p> <table border="1" data-bbox="532 737 1546 1142"> <tr><td>ANALYSER DEFIB QED-6H</td><td>FLUKE</td><td>QED 6</td></tr> <tr><td>ANALYSER DEFIB QED-6H</td><td>BIO TEK</td><td>QED 6H</td></tr> <tr><td>ANALYSER GAS FLW VT-PLUS</td><td>FLUKE</td><td>UT PLUS</td></tr> <tr><td>ANALYSER, ELECTRIC SAFETY</td><td>BC BIOMEDICAL</td><td>SA-2010S</td></tr> <tr><td>ANALYSER, ELECTRIC SAFETY</td><td>DATREND SYSTEMS</td><td>VPAD SAFETY ANALYZER</td></tr> <tr><td>ANALYSER, ELECTRICAL SAFETY</td><td>DATREND SYSTEMS</td><td>VPAD SAFETY ANALYZER</td></tr> <tr><td>ANALYZER DEFIBRILLATOR</td><td>BC GROUP</td><td>DA2006P</td></tr> <tr><td>ANALYZER, ELECTRICAL SAFETY</td><td>DATREND</td><td>VPAD-ES</td></tr> <tr><td>SIMULATOR MONITOR SYST</td><td>FLUKE</td><td>MPS 450</td></tr> <tr><td>SIMULATOR PATIENT ECG</td><td>BIO TEK</td><td>LH 3C</td></tr> <tr><td>SIMULATOR PULSE OXIMETER</td><td>FLUKE</td><td>INDEX 2 XLFE</td></tr> <tr><td>TESTER BP/NIBP BP PUMP 2</td><td>BC GROUP</td><td>NIBP1030KIT</td></tr> <tr><td>TESTER OXIMETER SP02</td><td>BC BIOMEDICAL</td><td>SPO-2000</td></tr> <tr><td>WATTMETER UPM-DT-1</td><td>BC BIOMEDICAL</td><td>USP-50SP</td></tr> </table>	ANALYSER DEFIB QED-6H	FLUKE	QED 6	ANALYSER DEFIB QED-6H	BIO TEK	QED 6H	ANALYSER GAS FLW VT-PLUS	FLUKE	UT PLUS	ANALYSER, ELECTRIC SAFETY	BC BIOMEDICAL	SA-2010S	ANALYSER, ELECTRIC SAFETY	DATREND SYSTEMS	VPAD SAFETY ANALYZER	ANALYSER, ELECTRICAL SAFETY	DATREND SYSTEMS	VPAD SAFETY ANALYZER	ANALYZER DEFIBRILLATOR	BC GROUP	DA2006P	ANALYZER, ELECTRICAL SAFETY	DATREND	VPAD-ES	SIMULATOR MONITOR SYST	FLUKE	MPS 450	SIMULATOR PATIENT ECG	BIO TEK	LH 3C	SIMULATOR PULSE OXIMETER	FLUKE	INDEX 2 XLFE	TESTER BP/NIBP BP PUMP 2	BC GROUP	NIBP1030KIT	TESTER OXIMETER SP02	BC BIOMEDICAL	SPO-2000	WATTMETER UPM-DT-1	BC BIOMEDICAL	USP-50SP
ANALYSER DEFIB QED-6H	FLUKE	QED 6																																									
ANALYSER DEFIB QED-6H	BIO TEK	QED 6H																																									
ANALYSER GAS FLW VT-PLUS	FLUKE	UT PLUS																																									
ANALYSER, ELECTRIC SAFETY	BC BIOMEDICAL	SA-2010S																																									
ANALYSER, ELECTRIC SAFETY	DATREND SYSTEMS	VPAD SAFETY ANALYZER																																									
ANALYSER, ELECTRICAL SAFETY	DATREND SYSTEMS	VPAD SAFETY ANALYZER																																									
ANALYZER DEFIBRILLATOR	BC GROUP	DA2006P																																									
ANALYZER, ELECTRICAL SAFETY	DATREND	VPAD-ES																																									
SIMULATOR MONITOR SYST	FLUKE	MPS 450																																									
SIMULATOR PATIENT ECG	BIO TEK	LH 3C																																									
SIMULATOR PULSE OXIMETER	FLUKE	INDEX 2 XLFE																																									
TESTER BP/NIBP BP PUMP 2	BC GROUP	NIBP1030KIT																																									
TESTER OXIMETER SP02	BC BIOMEDICAL	SPO-2000																																									
WATTMETER UPM-DT-1	BC BIOMEDICAL	USP-50SP																																									
<p><b>Answer 4:</b></p>	<p>The DND does not have a copy of the calibration procedures for the listed equipment.</p>																																										