



RETURN BIDS TO:

RETOURNER LES SOUMISSIONS À:

Public Works and Government Services / Travaux
publics et services gouvernementaux
Kingston Procurement
Des Acquisitions Kingston
86 Clarence Street, 2nd floor
Kingston
Ontario
K7L 1X3
Bid Fax: (613) 545-8067

**SOLICITATION AMENDMENT
MODIFICATION DE L'INVITATION**

The referenced document is hereby revised; unless otherwise
indicated, all other terms and conditions of the Solicitation
remain the same.

Ce document est par la présente révisé; sauf indication contraire,
les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

Vendor/Firm Name and Address
Raison sociale et adresse du
fournisseur/de l'entrepreneur

Issuing Office - Bureau de distribution
Public Works and Government Services / Travaux
publics et services gouvernementaux
Kingston Procurement
Des Acquisitions Kingston
86 Clarence Street, 2nd floor
Kingston
Ontario
K7L 1X3

Title - Sujet Microhardness Testing System	
Solicitation No. - N° de l'invitation W0046-195118/A	Amendment No. - N° modif. 001
Client Reference No. - N° de référence du client W0046-19-5118	Date 2019-01-25
GETS Reference No. - N° de référence de SEAG PW-\$KIN-555-7729	
File No. - N° de dossier KIN-8-50187 (555)	CCC No./N° CCC - FMS No./N° VME
Solicitation Closes - L'invitation prend fin at - à 02:00 PM on - le 2019-02-26	
Time Zone Fuseau horaire Eastern Standard Time EST	
F.O.B. - F.A.B. Specified Herein - Précisé dans les présentes Plant-Usine: <input type="checkbox"/> Destination: <input type="checkbox"/> Other-Autre: <input checked="" type="checkbox"/>	
Address Enquiries to: - Adresser toutes questions à: Webster, Sean	Buyer Id - Id de l'acheteur kin555
Telephone No. - N° de téléphone (613) 545-8843 ()	FAX No. - N° de FAX (613) 545-8067
Destination - of Goods, Services, and Construction: Destination - des biens, services et construction:	

Instructions: See Herein

Instructions: Voir aux présentes

Delivery Required - Livraison exigée	Delivery Offered - Livraison proposée
Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur	
Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur	
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date

Solicitation No. - N° de l'invitation
W0046-195118/A
Client Ref. No. - N° de réf. du client
W0046-19-5118

Amd. No. - N° de la modif.
001
File No. - N° du dossier
KIN-8-50187

Buyer ID - Id de l'acheteur
KIN555
CCC No./N° CCC - FMS No./N° VME

Amendment 001 – Stationary Collaborative Robot Assembly is being issued in response to Offeror's questions and Canada's responses.

Reference: Annex A, 3.1 Automatic Microindentation System Technical Specifications f) Automated Motorized Stage

- Q1: Given that the maximum size for a puck is 2" (50.8 mm) as specified in technical specification 3.1 (k) would it be possible to change technical specification 3.1 (f) to "Automated motorized stage, minimum 100x60mm"?
- A1: The 100x100mm (X-Y measurements) motorized stage would be considered a minimum size. Besides puck-mounted specimens, our lab runs tests on chunks of muzzle, breech, shell-casing and other types of fragments not mounted in pucks, so anything less than a 100x100mm stage would be considered inadequate for supporting such fragments.
-

ALL OTHER TERMS AND CONDITIONS REMAIN UNCHANGED