



RETURN BIDS TO:

RETOURNER LES SOUMISSIONS À:

Bid Receiving Public Works and Government
Services Canada/Réception des soumissions/Travaux
publics et Services gouvernementaux Canada
Building S-111, Rm C-114
101 Menin Rd. Garrison Petawawa
Petawawa
Ontario
K8H 2X3
Bid Fax: (613) 687-6656

**SOLICITATION AMENDMENT
MODIFICATION DE L'INVITATION**

The referenced document is hereby revised; unless otherwise
indicated, all other terms and conditions of the Solicitation
remain the same.

Ce document est par la présente révisé; sauf indication contraire,
les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

Address inquiries to the Contracting Authority at
wayne.cook@pwgsc-tpsgc.gc.ca

Vendor/Firm Name and Address

Raison sociale et adresse du
fournisseur/de l'entrepreneur

Issuing Office - Bureau de distribution

Public Works and Government Services Canada Supply
and Services Operation
Petawawa Procurement
Building S-111, Rm C-114
101 Menin Rd. Garrison Petawawa
Petawawa
Ontario
K8H 2X3

Title - Sujet deployable CT scanner	
Solicitation No. - N° de l'invitation W6369-18A019/B	Amendment No. - N° modif. 002
Client Reference No. - N° de référence du client W6369-18A019	Date 2019-03-04
GETS Reference No. - N° de référence de SEAG PW-\$PET-906-1555	
File No. - N° de dossier PET-8-49007 (906)	CCC No./N° CCC - FMS No./N° VME
Solicitation Closes - L'invitation prend fin at - à 02:00 PM on - le 2019-03-18	
Time Zone Fuseau horaire Eastern Daylight Saving Time EDT	
F.O.B. - F.A.B. Plant-Usine: <input checked="" type="checkbox"/> Destination: <input type="checkbox"/> Other-Autre: <input type="checkbox"/>	
Address Enquiries to: - Adresser toutes questions à: Cook, Wayne	Buyer Id - Id de l'acheteur pet906
Telephone No. - N° de téléphone (613) 401-0623 ()	FAX No. - N° de FAX (613) 687-6656
Destination - of Goods, Services, and Construction: Destination - des biens, services et construction:	

Instructions: See Herein

Instructions: Voir aux présentes

Delivery Required - Livraison exigée	Delivery Offered - Livraison proposée
Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur	
Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur	
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date

Amendement 002
Deployable CT Scanner

This amendment was raised to provide Canada's answers to Offeror's:

Question:

1) M34. States that the container when expanded, must have the following minimum external width of 5791 mm is this including the ramp

2) Is Annex D - to be left completely blank and will be completed at the time of Acceptance

3) Solicitation No. W6369-18A019/B Annex A Section 4 Standards and Specifications Item H

The CT Scanner Container must have all doors and openings sealed against the ingress of dust in accordance with **DEF STAN 07-55 Part 2 Section 4/1** Test D1. Proof of DEF- STAN certification must be provided prior to DND accepting delivery.

As per Addendum 3 from the previous **Solicitation** No. W6369-18A019/A 2018-08-27 the above standard was replaces with the following DEF Standard

As a minimum, the CT Scanner Container must have all doors and openings sealed against the ingress of airborne particles of fine matter in accordance with DEF-STAN 00-35 Part 3 / Issue 4 – Chapter 3-25 – Test CL25 – Wind Blown Dust and Sand.

Therefore can you please confirm which standard will be applicable for this Solicitation W6369-18A019/B

Answer:

1) The minimum expanded external width does not include the ramp.

2) Annex D will be completed during acceptance verification.

3) The Department of National Defence requires the CT Scanner Container to have all doors and openings sealed against the ingress of airborne particles of fine matter in accordance with DEF-STAN 00-35 Part 3 / Issue 4 – Chapter 3-25 – Test CL25 – Wind Blown Dust and Sand.