



## RETURN BIDS TO:

## RETOURNER LES SOUMISSIONS À:

Bid Receiving Public Works and Government  
Services Canada/Réception des soumissions Travaux  
publics et Services gouvernementaux Canada  
Government of Canada Building  
101 - 22nd Street East, Suite 110  
Saskatoon  
Saskatchewan  
S7K 0E1  
Bid Fax: (306) 975-5397

## SOLICITATION AMENDMENT MODIFICATION DE L'INVITATION

The referenced document is hereby revised; unless otherwise  
indicated, all other terms and conditions of the Solicitation  
remain the same.

Ce document est par la présente révisé; sauf indication contraire,  
les modalités de l'invitation demeurent les mêmes.

### Comments - Commentaires

### Vendor/Firm Name and Address

Raison sociale et adresse du  
fournisseur/de l'entrepreneur

### Issuing Office - Bureau de distribution

Public Works and Government Services Canada/Travaux  
publics et Services gouvernementaux Canada  
Harry Hays Building (HHB)  
Room 759, 220-4th Avenue SE  
Calgary  
Alberta  
T2G 4X3

|   |  |
|---|--|
| <b>Title - Sujet</b><br>Liquid Chromatography System  |  |
| <b>Solicitation No. - N° de l'invitation</b><br>01581-200445/A  | <b>Amendment No. - N° modif.</b><br>001      |
| <b>Client Reference No. - N° de référence du client</b><br>01581-200445   | <b>Date</b><br>2019-11-14                    |
| <b>GETS Reference No. - N° de référence de SEAG</b><br>PW-SCAL-137-6980   |  |
| <b>File No. - N° de dossier</b><br>CAL-9-42082 (137)  | <b>CCC No./N° CCC - FMS No./N° VME</b>       |
| <b>Solicitation Closes - L'invitation prend fin</b><br><b>at - à 02:00 PM</b><br><b>on - le 2019-12-09</b>  |  |
| <b>Time Zone</b><br><b>Fuseau horaire</b><br>Central Standard Time<br>CST   |  |
| <b>F.O.B. - F.A.B.</b><br><b>Plant-Usine:</b> <input type="checkbox"/> <b>Destination:</b> <input checked="" type="checkbox"/> <b>Other-Autre:</b> <input type="checkbox"/>   |  |
| <b>Address Enquiries to: - Adresser toutes questions à:</b><br>Loi, Ngan  | <b>Buyer Id - Id de l'acheteur</b><br>cal137 |
| <b>Telephone No. - N° de téléphone</b><br>(403) 973-2796 ( )  | <b>FAX No. - N° de FAX</b><br>(306) 975-5397 |
| <b>Destination - of Goods, Services, and Construction:</b><br><b>Destination - des biens, services et construction:</b><br>Swift Current Research and Development Centre<br>1 Airport Road<br>Swift Current, Saskatchewan S9H 3X2<br>Canada |  |

Instructions: See Herein

Instructions: Voir aux présentes

|  |  |
|--|--|
| <b>Delivery Required - Livraison exigée</b>  | <b>Delivery Offered - Livraison proposée</b> |
| <b>Vendor/Firm Name and Address</b><br><b>Raison sociale et adresse du fournisseur/de l'entrepreneur</b>   |  |
| <b>Telephone No. - N° de téléphone</b><br><b>Facsimile No. - N° de télécopieur</b>   |  |
| <b>Name and title of person authorized to sign on behalf of Vendor/Firm</b><br><b>(type or print)</b><br><b>Nom et titre de la personne autorisée à signer au nom du fournisseur/</b><br><b>de l'entrepreneur (taper ou écrire en caractères d'imprimerie)</b> |  |
| <b>Signature</b>   | <b>Date</b>                                  |

---

This Amendment 001 is being raised to provide answers to the questions that have been received and incorporate the following information into Solicitation 01581-200445/A:

**Question 1:** Could you please clarify specification #8 "Instrument must support a minimum of two columns running in parallel, with independent temperature control". Are you looking for an HPLC with a column-switching valve that will allow you to switch between two columns (independently heated) for different analysis? If not, could you please elaborate on your requirement?

**Answer 1:** It would be a valve that allows for quickly switching between two or more columns, which are independently heated, for different analyses or sample requirements.

**Question 2:** Is independent column temperature control mandatory or preferred? Do you plan to run the two columns at different temperatures or the same temperature when ran in parallel?

**Answer 2:** Independent temperature control for each column is mandatory. We need to have the option of running each column at an independent temperature to be sure we can accommodate a variety of methods.

**1. Page 13 of 15, Annex A – Requirement, Compliance Matrix:**

**Remove:** Item #8 Instrument must support a minimum of two columns running in parallel, with independent temperature control.

**Replace with:** Item #8 Instrument must support a minimum of two columns installed in parallel with independent temperature control in order to allow for quick transition between methods.

**All other terms and conditions remain the same**