



**RETURN BIDS TO:**

**RETOURNER LES SOUMISSIONS À:**

Public Works and Government Services / Travaux  
publics et services gouvernementaux  
Kingston Procurement  
Des Acquisitions Kingston  
86 Clarence Street, 2nd floor  
Kingston  
Ontario  
K7L 1X3  
Bid Fax: (613) 545-8067

**SOLICITATION AMENDMENT  
MODIFICATION DE L'INVITATION**

The referenced document is hereby revised; unless otherwise  
indicated, all other terms and conditions of the Solicitation  
remain the same.

Ce document est par la présente révisé; sauf indication contraire,  
les modalités de l'invitation demeurent les mêmes.

**Comments - Commentaires**

Samples and standards must be requested from Mike  
Littlefield (mike.littlefield@pwgsc.gc.ca ) by  
NOVEMBER 6th, 2019 to perform these tests.

**Vendor/Firm Name and Address**

Raison sociale et adresse du  
fournisseur/de l'entrepreneur

**Issuing Office - Bureau de distribution**

Public Works and Government Services / Travaux  
publics et services gouvernementaux  
Kingston Procurement  
Des Acquisitions Kingston  
86 Clarence Street, 2nd floor  
Kingston  
Ontario  
K7L 1X3

<b>Title - Sujet</b> Chromatography Spectrometer	
<b>Solicitation No. - N° de l'invitation</b> KW405-190288/B	<b>Amendment No. - N° modif.</b> 003
<b>Client Reference No. - N° de référence du client</b> 500039959	<b>Date</b> 2019-11-18
<b>GETS Reference No. - N° de référence de SEAG</b> PW-\$KIN-930-7919	
<b>File No. - N° de dossier</b> KIN-8-50049 (930)	<b>CCC No./N° CCC - FMS No./N° VME</b>
<b>Solicitation Closes - L'invitation prend fin</b> <b>at - à 02:00 PM</b> <b>on - le 2019-11-27</b>	
<b>Time Zone</b> <b>Fuseau horaire</b> Eastern Standard Time EST	
<b>F.O.B. - F.A.B.</b> <b>Plant-Usine:</b> <input type="checkbox"/> <b>Destination:</b> <input checked="" type="checkbox"/> <b>Other-Autre:</b> <input type="checkbox"/>	
<b>Address Enquiries to: - Adresser toutes questions à:</b> Littlefield, Mike	<b>Buyer Id - Id de l'acheteur</b> kin930
<b>Telephone No. - N° de téléphone</b> (613) 545-8058 ( )	<b>FAX No. - N° de FAX</b> (613) 545-8067
<b>Destination - of Goods, Services, and Construction:</b> <b>Destination - des biens, services et construction:</b> Environment and Climate Change Canada(ECCC) 867 Lakeshore Rd Burlington, Ontario L7S 1A1	

Instructions: See Herein

Instructions: Voir aux présentes

<b>Delivery Required - Livraison exigée</b>	<b>Delivery Offered - Livraison proposée</b>
<b>Vendor/Firm Name and Address</b> <b>Raison sociale et adresse du fournisseur/de l'entrepreneur</b>	
<b>Telephone No. - N° de téléphone</b> <b>Facsimile No. - N° de télécopieur</b>	
<b>Name and title of person authorized to sign on behalf of Vendor/Firm</b> <b>(type or print)</b> <b>Nom et titre de la personne autorisée à signer au nom du fournisseur/</b> <b>de l'entrepreneur (taper ou écrire en caractères d'imprimerie)</b>	
<b>Signature</b>	<b>Date</b>

---

**Amendment 003** – is being issued in response to Offeror's questions and Canada's responses.

---

**Q1:** *Within the RFP requirements there is no mention of a Nitrogen Generator. Nitrogen will be required for operation of the LCMSMS systems.  
Are you able to confirm that your client has an available Nitrogen supply for use with the LCMSMS systems and does not need vendors to include a Nitrogen generator(s) within their proposal?*

**A1:** We currently have a Nitrogen Generator that will be able to supply the new LC/QQQ systems.

---

**Q2:** *Will you accept a PC processor with Intel Core i5-8500 @ 3.00 GHz?*

**A2:** Yes.

**Reference:** **Section 4- Operating/Data Computer System Hardware, subsection 4.1**

**Delete:** 4.1 The PC must at least be an i7 4 type processor operating at 3.2 GHz or Better. Must be able to simultaneous data acquisition and backup (RAID array)

**Insert:** 4.1 The PC must at least be an i5 type processor operating at 3.0 GHz or Better. Must be able to simultaneous data acquisition and backup (RAID array)

**Reference:** **Annex C, Section 4 – MT 4.1**

**Delete:** **MT 4.1** The PC must at least be an i7 4 type processor operating at 3.2 GHz or Better. Must be able to simultaneous data acquisition and backup (RAID array)

**Insert:** **MT 4.1** The PC must at least be an i5 type processor operating at 3.0 GHz or Better. Must be able to simultaneous data acquisition and backup (RAID array)

---

**Q3:** *Will you accept a UHD 630 onboard video card that uses the computer RAM?*

**A3:** Yes, an onboard video card will meet our requirement.

**Reference:** **Section 4- Operating/Data Computer System Hardware, subsection 4.2**

**Delete:** 4.2 The desktop must have a graphic board with 2 GB RAM or better.

**Insert:** 4.2 The desktop must have a graphic board with 2 GB RAM or onboard video card.

**Reference:** **Annex C, Section 4 – MT 4.2**

**Delete:** **MT 4.2** The desktop must have a graphic board with 2 GB RAM or better.

**Insert:** **MT 4.2** The desktop must have a graphic board with 2 GB RAM or onboard video card.

---

---

**Q4:** *Will you accept an 8x DVD+/-RW 9.5mm Optical Disk Drive?*

**A4:** Yes this acceptable, a reduction in write speed should not affect our operating needs.

**Reference:** **Section 4- Operating/Data Computer System Hardware, subsection 4.5**

**Delete:** 4.5 16x (or greater) DVD-/+r RW, Drive

**Insert:** 4.5 8x (or greater) DVD-/+r RW, Drive

**Reference:** **Annex C, Section 4 – MT 4.5**

**Delete:** **MT 4.5** 16x (or greater) DVD-/+r RW, Drive

**Insert:** **MT 4.5** 8x (or greater) DVD-/+r RW, Drive

---

**Q5:** *Will our bid be automatically disqualified if we submit requested changes to the terms of the resultant contract?*

**A5:** Yes. Your company's bid would be disqualified as the Government of Canada deems any proposed changes to existing terms and conditions through a competitive process as a counter offer.

---

**ALL OTHER TERMS AND CONDITIONS REMAIN UNCHANGED**