



RETURN BIDS TO:

RETOURNER LES SOUMISSIONS À:

Public Works and Government Services / Travaux
publics et services gouvernementaux
Kingston Procurement
Des Acquisitions Kingston
86 Clarence Street, 2nd floor
Kingston
Ontario
K7L 1X3
Bid Fax: (613) 545-8067

**SOLICITATION AMENDMENT
MODIFICATION DE L'INVITATION**

The referenced document is hereby revised; unless otherwise
indicated, all other terms and conditions of the Solicitation
remain the same.

Ce document est par la présente révisé; sauf indication contraire,
les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

Address all inquiries to:
Tammy.Weaver@tpsgc-pwgsc.gc.ca

Vendor/Firm Name and Address

Raison sociale et adresse du
fournisseur/de l'entrepreneur

Issuing Office - Bureau de distribution

Public Works and Government Services / Travaux
publics et services gouvernementaux
Kingston Procurement
Des Acquisitions Kingston
86 Clarence Street, 2nd floor
Kingston
Ontario
K7L 1X3

Title - Sujet Insulation Resistivity Tester	
Solicitation No. - N° de l'invitation 23240-200399/A	Amendment No. - N° modif. 002
Client Reference No. - N° de référence du client 23240-200399	Date 2019-12-20
GETS Reference No. - N° de référence de SEAG PW-\$KIN-508-7942	
File No. - N° de dossier KIN-9-52131 (508)	CCC No./N° CCC - FMS No./N° VME
Solicitation Closes - L'invitation prend fin at - à 02:00 PM on - le 2020-01-06	
Time Zone Fuseau horaire Eastern Standard Time EST	
F.O.B. - F.A.B. Plant-Usine: <input type="checkbox"/> Destination: <input type="checkbox"/> Other-Autre: <input type="checkbox"/>	
Address Enquiries to: - Adresser toutes questions à: Weaver, Tammy	Buyer Id - Id de l'acheteur kin508
Telephone No. - N° de téléphone (613) 484-1809 ()	FAX No. - N° de FAX (613) 545-8067
Destination - of Goods, Services, and Construction: Destination - des biens, services et construction:	

Instructions: See Herein

Instructions: Voir aux présentes

Delivery Required - Livraison exigée	Delivery Offered - Livraison proposée
Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur	
Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur	
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date

The following questions have been submitted by a potential Bidder:

Regarding the subject solicitation would you please clarify following questions:

1. With reference to Annex A, item A1-3.

We are offering from a well known European manufacturer. Please confirm equivalent European fire protection and safety regulation (machinery directive standard "DIRECTIVE 2006/42/EC") is also acceptable. According to this regulation Machinery must be designed and constructed in such a way as to avoid any risk of fire or overheating posed by the machinery itself or by gases, liquids, dust, vapours or other substances produced or used by the machinery.

We are required to meet the standards as indicated in the contract to ensure the safety of our employees and to meet federal regulations. We are not experts on the standards themselves and we are not experts on the European standards. We would accept a document stamped by an engineer licensed in Ontario stating that the equipment meets the standards, or that the DIRECTIVE 2006/42/EC is equivalent.

(We may be able to provide contact information to local engineers who can answer the question if it is acceptable to public works.)

2. Would you please clarify if we have to include a laptop in our offer or this is something NRC will provide?

The offer must include a laptop with the installed software to run the equipment.

3. Would you please confirm that NRC is responsible for onsite Compressed air supply?

CanmetMATERIALS (Natural Resources Canada) will be responsible for onsite compressed air supply.

4. With reference to Annex A, item A2-5 "A Hydraulic Press with a capacity of 2,000 lb [10,000 N], and with mountings to accommodate the test head, testing specimens or punchings".

Our system is driven by pneumatic cylinder. The requested force is 2N/mm² for the total contact area of 645 mm². please confirm this is acceptable.

The equipment must comply to ASTM standard A717/A717M. The standard specified a Hydraulic Press. Thus, only Hydraulic Press is acceptable.

The requested force of 2,000 lb [10,000 N] is for the total contact area of 645 mm².