



RETURN BIDS TO:

RETOURNER LES SOUMISSIONS À:

Public Works and Government Services / Travaux
publics et services gouvernementaux
Kingston Procurement
Des Acquisitions Kingston
86 Clarence Street, 2nd floor
Kingston
Ontario
K7L 1X3
Bid Fax: (613) 545-8067

**SOLICITATION AMENDMENT
MODIFICATION DE L'INVITATION**

The referenced document is hereby revised; unless otherwise
indicated, all other terms and conditions of the Solicitation
remain the same.

Ce document est par la présente révisé; sauf indication contraire,
les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

Vendor/Firm Name and Address
Raison sociale et adresse du
fournisseur/de l'entrepreneur

Issuing Office - Bureau de distribution
Public Works and Government Services / Travaux
publics et services gouvernementaux
Kingston Procurement
Des Acquisitions Kingston
86 Clarence Street, 2nd floor
Kingston
Ontario
K7L 1X3

Title - Sujet Ultrasound System	
Solicitation No. - N° de l'invitation 21401-190002/A	Amendment No. - N° modif. 002
Client Reference No. - N° de référence du client 21401-190002	Date 2019-12-24
GETS Reference No. - N° de référence de SEAG PW-\$KIN-630-7946	
File No. - N° de dossier KIN-9-52127 (630)	CCC No./N° CCC - FMS No./N° VME
Solicitation Closes - L'invitation prend fin at - à 02:00 PM on - le 2020-01-20	
Time Zone Fuseau horaire Eastern Standard Time EST	
F.O.B. - F.A.B. Plant-Usine: <input type="checkbox"/> Destination: <input type="checkbox"/> Other-Autre: <input type="checkbox"/>	
Address Enquiries to: - Adresser toutes questions à: Holt, Judy	Buyer Id - Id de l'acheteur kin630
Telephone No. - N° de téléphone (613) 536-4995 ()	FAX No. - N° de FAX (613) 545-8067
Destination - of Goods, Services, and Construction: Destination - des biens, services et construction: See herein	

Instructions: See Herein

Instructions: Voir aux présentes

Delivery Required - Livraison exigée	Delivery Offered - Livraison proposée
Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur	
Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur	
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date

Solicitation No. - N° de l'invitation
21401-190002/A
Client Ref. No. - N° de réf. du client

Amd. No. - N° de la modif.
002
File No. - N° du dossier
KIN-9-52127

Buyer ID - Id de l'acheteur
KIN630
CCC No./N° CCC - FMS No./N° VME

This amendment is issued to Solicitation number 21401-190002/A to answer the following questions an

A. Questions from Industry and answers from Canada:

Question - Section 2: Minimum mandatory spec, item 2.0 - Is the OB/Gyn optional for now? If needed later it is just a license to turn on the OB/Gyn application.

Response – As per specification 2.0.1, the system must be capable of upgrading software and transducers to be used for OB/Gyn examinations. This upgrading may be required at a later date.