



RETURN BIDS TO:

RETOURNER LES SOUMISSIONS À:

Public Works and Government Services Canada
Canada Place/Place du Canada
10th Floor/10e étage
9700 Jasper Ave/9700 ave Jasper
Edmonton
Alberta
T5J 4C3
Bid Fax: (780) 497-3510

**SOLICITATION AMENDMENT
MODIFICATION DE L'INVITATION**

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Solicitation remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

Vendor/Firm Name and Address
Raison sociale et adresse du
fournisseur/de l'entrepreneur

Issuing Office - Bureau de distribution
Public Works and Government Services Canada
Canada Place/Place du Canada
10th Floor/10e étage
9700 Jasper Ave/9700 ave Jasper
Edmonton
Alberta
T5J 4C3

Title - Sujet Sample Extract Parallel Evaporator	
Solicitation No. - N° de l'invitation 01634-200649/A	Amendment No. - N° modif. 001
Client Reference No. - N° de référence du client 01634-200649	Date 2020-01-06
GETS Reference No. - N° de référence de SEAG PW-\$EDM-034-11754	
File No. - N° de dossier EDM-9-42173 (034)	CCC No./N° CCC - FMS No./N° VME
Solicitation Closes - L'invitation prend fin at - à 02:00 PM on - le 2020-01-21	Time Zone Fuseau horaire Mountain Standard Time MST
F.O.B. - F.A.B. Plant-Usine: <input type="checkbox"/> Destination: <input type="checkbox"/> Other-Autre: <input type="checkbox"/>	
Address Enquiries to: - Adresser toutes questions à: Nigam, Nidhi	Buyer Id - Id de l'acheteur edm034
Telephone No. - N° de téléphone (587) 532-8142 ()	FAX No. - N° de FAX (780) 497-3510
Destination - of Goods, Services, and Construction: Destination - des biens, services et construction:	

Instructions: See Herein

Instructions: Voir aux présentes

Delivery Required - Livraison exigée	Delivery Offered - Livraison proposée
Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur	
Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur	
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date

This amendment has been raised to answer the questions received from the industry.

Q1. Regarding mandatory specification item #3 are you willing to accept parallel evaporation of up to 18 samples vs the 24 listed?

R1. Yes parallel evaporation of up to 18 samples is acceptable.
Please see revised Annex B, minimum mandatory performance specifications.

Delete in its entirety: Annex B, minimum mandatory performance specifications.

Insert:

ANNEX ``B``

Minimum Mandatory Performance Specifications

Compliance Matrix

A complete list of the minimum mandatory performance specifications are detailed below in the "Compliance Matrix". Bidders are to clearly demonstrate compliance with each mandatory specification.

1. Bidders **to** show compliance by addressing each performance specification in the Compliance Matrix, whether the product offered "meets" or "doesn't meet".
2. Bidders are requested to indicate how they meet each performance specification by recording this information under the Performance Specification Offered column in the Compliance Matrix.
3. It is requested that supporting technical documentation, including but not limited to, specification sheets, technical brochures, photographs or illustrations be provided with the bid at solicitation close and be cross-referenced on the Compliance Matrix for each performance specification to outline where in the supporting technical documentation it demonstrates compliance. It is the Bidders responsibility to ensure that the submitted supporting technical documentation provides detail to prove that the proposed product(s) meet the requirements of the Performance Specification. If published supporting technical document is not available, the Bidder should prepare a written narrative complete with a detailed explanation of how its bid demonstrates technical compliance.
4. If the supporting documentation referenced above has not been provided at bid closing, the Contracting Authority will notify the Bidder that they must provide supporting documentation within two (2) business days following notification. Failure to comply with the request of the Contracting Authority within that time period, will deem the bid non-responsive and the bid will be given no further consideration.
5. Bidders must address any concerns with the performance specifications in written detail to the Contracting Authority before bid closing as outlined in the Request for Proposal (RFP) document.
6. Failure to meet each mandatory performance specification will result in the bid being deemed non-responsive, and be given no further consideration.

COMPLIANCE MATRIX – MINIMUM MANDATORY PERFORMANCE SPECIFICATIONS:

Item #	Mandatory Performance Specification	Performance Specification Met?		Performance Specification Offered: Bidder should indicate how they meet the performance specification by recording this information in this column	Cross Reference: In this column, Bidders should cross-reference where this performance specification is indicated in their supporting documents.
		Yes	No		
1	Sample extract parallel evaporator must be able to concentrate and evaporate multiple samples to a predefined volume (i.e., 25 ml to 0 mL and 25 mL to 0.3 mL). Suitable glassware should be supplied if regular glassware is not compatible with the unit.				
2	Evaporation and concentration process should be automated with minimal user input				
3	Unit must support parallel evaporation of a minimum of 18 samples				
4	Ability to recover the solvent from the evaporation process				
5	During evaporation and concentration, sample extract should not bump or dry on the sample container walls (to avoid cross contamination of samples and achieve higher recoveries).				
6	Sound Level: Must operate with sound levels within acceptable range, < 70 dB				
7	Power requirements must be suitable for North American standards.				

Solicitation No. - N° de l'invitation
01634-200649/A
Client Ref. No. - N° de réf. du client
01634-200649

Amd. No. - N° de la modif.
001
File No. - N° du dossier
EDM-9-42173

Buyer ID - Id de l'acheteur
Edm034
CCC No. /N° CCC - FMS No./N° VME

ALL OTER TERMS AND CONDITIONS REMAIN THE SAME AND IN FULL EFFECT.