



**SOLICITATION AMENDMENT/
MODIFICATION DE L'INVITATION**

**RETURN BIDS TO /
RETOURNER LES SOUMISSIONS À:**

By e-mail to: binh.duong@forces.gc.ca
Director Services Contracting (D Svcs C) 3-2-8-2
Att: Binh Duong

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Solicitation remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de l'invitation demeurent les mêmes.

Title / Titre: Large Volume Infusion Pumps (LVIP)	
Solicitation No / No de l'invitation: W6369-20-A043	Amendment No. - N° de la modification 002
Date of Amendment – Date de la modification 14 January 2020	
Address Enquiries to – Adresser toutes questions à: Binh Duong Director of Services Contracting (D Svcs C) 3 101 Colonel By Drive Ottawa, Ontario K1A 0K2 binh.duong@forces.gc.ca	
Telephone No. / N° de téléphone: 819-939-8539	FAX No / No de fax:
Destination: See Herein	

Instructions:

Municipal taxes are not applicable. Unless otherwise specified herein all prices quoted must include all applicable Canadian customs duties, GST/HST, excise taxes and are to be delivered Delivery Duty Paid including all delivery charges to destination(s) as indicated. The amount of the Goods and Services Tax/Harmonized Sales Tax is to be shown as a separate item.

Instructions:

Les taxes municipales ne s'appliquent pas. Sauf indication contraire, les prix indiqués doivent comprendre les droits de douane canadiens, la TPS/TVH et la taxe d'accise. Les biens doivent être livrés «rendu droits acquittés», tous frais de livraison compris, à la ou aux destinations indiquées. Le montant de la taxe sur les produits et services/taxe de vente harmonisée doit

**Solicitation Closes /
L'invitation prend fin:**

At / à :
2:00 PM
Eastern Daylight Time (EDT)

On / le : 23 January 2020

Delivery required / Livraison exigée: 30 March 2020 or as soon as possible.	Delivery offered / Livraison proposée:
Vendor Name and Address / Raison sociale et adresse du fournisseur:	
Name and title of person authorized to sign on behalf of vendor (type or print) / Nom et titre de la personne autorisée à signer au nom du fournisseur (caractère d'imprimerie):	
Name / Nom: _____	Title / Titre: _____
Signature: _____	Date: _____



AMENDMENT 002 TO SOLICITATION NUMBER W6369-20-A043 IS RAISED TO:

1. Provide clarification and answers to questions from potential suppliers.
2. Update Appendix 1 to Annex A – Requirement section 5.2.
3. Extend the solicitation closing date by three (3) days from 20 January 2020 to 23 January 2020.

QUESTIONS AND ANSWERS:

Question 1	<p>* On page 34, section 6 iii, Two (2) general purpose infusion sets per infusion pump model per device...</p> <p>* Can you please clarify if the 2 infusion sets are for the pump with 2 modules or is it 2 infusion sets per module?</p> <p>* If it is the latter, based on the piggyback infusion pre-requisites listed within the RFP, Are vendors expected to supply a primary and secondary infusion set?</p>
Answer 1	<p>This requirement is discussing accessories that are provided by the company. The company must provide 2 general purpose infusion sets per module (therefore 4 general purpose infusion sets with each device, no secondary sets are required).</p>

Question 2	<p>* On page 32, section 5.2 viii, it is stated that the pump “permit secondary infusion to resume automatically after the primary infusion has completed, also known as Automatic Piggybacking.”</p> <p>* Can you please clarify if this was written in error and is in fact the opposite where the primary infusion would resume automatically after the secondary infusion has completed?</p>
Answer 2	<p>Yes, the primary infusion should resume after the secondary line has finished. Please See new Appendix 1 to Annex A – Requirement, Article 5.2.</p>

Question 3	<p>Page 22 indicates that deliverables must be received by 30 March 2020 or as soon as possible, whereas page 28 indicates that delivery must be by 28 February 2020 or as soon as possible. Given the difference in dates, Should vendors assume that delivery is required as soon as possible after contract signing?</p>
Answer 3	<p>The Large Volume Infusion Pumps (LVIP) must be delivered by 28 February 2020 or as soon as possible; and the training sessions must be delivered within thirty (30) calendar days of delivery of the initial LVIP material (by 30 March 2020) or as soon as possible.</p>

Question 4	<p>Are vendors allowed to reproduce the pricing appendix into an Excel spreadsheet and provide additional comments?</p>
Answer 4	<p>The Pricing Schedule must be submitted as per the RFP. Please refer to section 3.3 Section II Financial Bid: Bidders must submit their financial bid in accordance with the</p>



National Defence

Défense nationale

National Defence Headquarters
Ottawa, Ontario
K1A 0K2

Quartier général de la Défense nationale
Ottawa (Ontario)
K1A 0K2

	Pricing Schedule detailed in Attachment 1 to Part 3. If Bidders add any conditions or makes changes to the Pricing Schedule, the Bidder's financial bid will be declared non-responsive.
--	--

Question 5	Based on the RFP requirement, we believe The Department of Defense may be require pharmacy support for PCA and syringe drug library build during the option period: Would it be possible to add a Pharmacy Support service category related to the option period in the pricing proposal?
Answer 5	No, pharmacy support for PCA is not required.

SOLICITATION NUMBER W6369-20-A043 IS HEREBY AMENDED AS FOLLOWS:

1. DELETE Appendix 1 to Annex A – Requirement, Article 5.2 in its entirety, and **REPLACE** with the following:

5.2 Technical – Infusion Pump Module

- i. Minimum flow rate range of 0.1 to 999 milliliters per hour.
- ii. Depending on the flow rate, minimum rate increments of 0.1, 0.5, 1.0 milliliters per hour.
- iii. Minimum Keep Vein Open (KVO) rate of 0.5 – 5 milliliters per hour.
- iv. Maximum rate range accuracy of ± five (5) percent (%) for flow rates between 1 to 999 milliliters per hour.
- v. Minimum Volume to be Infused range of one (1) – 6000 milliliters.
- vi. Minimum pressure range of five (5) to ten (10) pounds per square inch (psi).
- vii. Accept fluids of differing densities and consistencies ranging from normal saline to blood products.
- viii. Permit a primary infusion to resume automatically after the secondary infusion has completed, also known as Automatic Piggybacking.

2. DELETE from Page 1, “20 January 2020” and **REPLACE** with “23 January 2020”.

ALL OTHER TERMS AND CONDITIONS REMAIN THE SAME.