



RETURN BIDS TO:

RETOURNER LES SOUMISSIONS À:

Public Works and Government Services Canada
Canada Place/Place du Canada
10th Floor/10e étage
9700 Jasper Ave/9700 ave Jasper
Edmonton
Alberta
T5J 4C3
Bid Fax: (780) 497-3510

**SOLICITATION AMENDMENT
MODIFICATION DE L'INVITATION**

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Solicitation remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

Vendor/Firm Name and Address
Raison sociale et adresse du
fournisseur/de l'entrepreneur

Issuing Office - Bureau de distribution
Public Works and Government Services Canada
Canada Place/Place du Canada
10th Floor/10e étage
9700 Jasper Ave/9700 ave Jasper
Edmonton
Alberta
T5J 4C3

Title - Sujet HTRF Plate Reader	
Solicitation No. - N° de l'invitation W7714-207318/A	Amendment No. - N° modif. 001
Client Reference No. - N° de référence du client W7714-207318	Date 2020-01-27
GETS Reference No. - N° de référence de SEAG PW-\$EDM-024-11766	
File No. - N° de dossier EDM-9-42187 (024)	CCC No./N° CCC - FMS No./N° VME
Solicitation Closes - L'invitation prend fin at - à 02:00 PM on - le 2020-02-06	Time Zone Fuseau horaire Mountain Standard Time MST
F.O.B. - F.A.B. Specified Herein - Précisé dans les présentes Plant-Usine: <input type="checkbox"/> Destination: <input type="checkbox"/> Other-Autre: <input checked="" type="checkbox"/>	
Address Enquiries to: - Adresser toutes questions à: Tiet, Anthony	Buyer Id - Id de l'acheteur edm024
Telephone No. - N° de téléphone (587) 926-1376 ()	FAX No. - N° de FAX (780) 497-3510
Destination - of Goods, Services, and Construction: Destination - des biens, services et construction:	

Instructions: See Herein

Instructions: Voir aux présentes

Delivery Required - Livraison exigée	Delivery Offered - Livraison proposée
Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur	
Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur	
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date

Amendment 001 has been raised to answer the following questions and make the following changes:

REVISED CLOSING DATE OF:

02:00 PM MST on 2020-02-06

- Question 1:** In Annex "B", Compliance Matrix – Minimum Mandatory Performance Specifications, item #1, will you accept a 616nm emission filter as the HTRF filter?
- Answer 1:** Yes, Annex "B", Compliance Matrix – Minimum Mandatory Performance Specifications, item #1 has been amended.
- Question 2:** In Annex "B", Compliance Matrix – Minimum Mandatory Performance Specifications, item #1, we are offering an HTRF Plate Reader with monochrometer (scanning) HTRF performance. This method offers more flexibility for HTRF compared to a filter-based system as any nm combination can be added rather than being restricted by filter sets and offers higher performance than filter-based technology. The offered system has auto-cut off filters built into the system to reduce bleed through with narrow stoke shifts. Would you please confirm this is acceptable?
- Answer 2:** Yes, Annex "B", Compliance Matrix – Minimum Mandatory Performance Specifications, item #1 has been amended.
- Question 3:** In Annex "B", Compliance Matrix – Minimum Mandatory Performance Specifications, item #3, please clarify if GLP/GMP is required?
- Answer 3:** Software containing measures for GLP-compliant is required, the combination of GLP/GMP is acceptable.
- Question 4:** In Annex "B", Compliance Matrix – Minimum Mandatory Performance Specifications, item #3, please clarify if GLP/SMP is required?
- Answer 4:** Software containing measures for GLP-compliant is required, the combination of GLP/SMP is acceptable.

On page 14 of 16, under **COMPLIANCE MATRIX – MINIMUM MANDATORY PERFORMANCE SPECIFICATIONS:**

DELETE: In its entirety

INSERT:

Requirement:	Manufacturer(s) Offered:	Model Number(s) Offered:
Homogenous Time Resolved Fluorescence (HTRF) Plate Reader		

Item #	Performance Specification	Status (M) Mandatory (D) Desirable*	Performance Specification Met? Indicate either Yes/No	Performance Specification Offered: Bidder <u>should</u> indicate how they meet the performance specification by recording this information in this column	Cross Reference: In this column, Bidders should cross-reference where this performance specification is indicated in their supporting documents.
1	Top reading HTRF with ability to have excitation at 340nm and emission at either 616nm or 620nm and 665nm.	M			
2	Capability for top and bottom reading of absorbance and fluorescence either by monochromator or combination with filter sets, and an additional luminescence reading package. Detection methods must include fluorescence intensity/ FRET, unfiltered luminescence, filtered luminescence, fluorescence polarization.	M			
3	Software containing measures for GLP-compliant data retention (software must have upgrade for FDA 21 CFR Part 11 compliance).	M			

ALL OTHER TERMS AND CONDITIONS REMAIN THE SAME AND IN FULL EFFECT.