



RETURN BIDS TO:

RETOURNER LES SOUMISSIONS À:

Bid Receiving - PWGSC / Réception des soumissions -
TPSGC

11 Laurier St./ 11 rue, Laurier

Place du Portage, Phase III

Core 0B2 / Noyau 0B2

Gatineau, Québec K1A 0S5

Bid Fax: (819) 997-9776

**SOLICITATION AMENDMENT
MODIFICATION DE L'INVITATION**

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Solicitation remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

Vendor/Firm Name and Address

Raison sociale et adresse du
fournisseur/de l'entrepreneur

Issuing Office - Bureau de distribution

Scientific, Medical and Photographic Division /
Division de l'équipement scientifique, des produits
photographiques et pharmaceutiques
L'Esplanade Laurier
140 O'Connor Street,
East Tower, 7th Floor
Ottawa
Ontario
K1A 0S5

Title - Sujet AED	
Solicitation No. - N° de l'invitation 31184-207568/A	Amendment No. - N° modif. 003
Client Reference No. - N° de référence du client 31184-207568	Date 2020-03-17
GETS Reference No. - N° de référence de SEAG PW-\$\$\$PV-956-78451	
File No. - N° de dossier pv956.31184-207568	CCC No./N° CCC - FMS No./N° VME
Solicitation Closes - L'invitation prend fin at - à 02:00 PM on - le 2020-03-23	
Time Zone Fuseau horaire Eastern Daylight Saving Time EDT	
F.O.B. - F.A.B. Plant-Usine: <input type="checkbox"/> Destination: <input checked="" type="checkbox"/> Other-Autre: <input type="checkbox"/>	
Address Enquiries to: - Adresser toutes questions à: Courteau, Robert	Buyer Id - Id de l'acheteur pv956
Telephone No. - N° de téléphone (343) 550-1614 ()	FAX No. - N° de FAX () -
Destination - of Goods, Services, and Construction: Destination - des biens, services et construction:	

Instructions: See Herein

Instructions: Voir aux présentes

Delivery Required - Livraison exigée	Delivery Offered - Livraison proposée
Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur	
Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur	
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date

Amendment 003 is raised to answer bidder questions:

1. Mandatory Requirement: "1 x cartridge that combines battery and electrode pads for the AED".
As there is currently only one device on the market that has this capability, we ask that this requirement be changed to the battery and electrode much have the same expiration length to be replaced at the same time to allow more than one device to be proposed and ensure you receive the best unit for your needs.

Answer: No

2. Mandatory Requirement: "The AED unit must be Semi-Automatic or fully Automatic in operation."
Can you please clarify if you would like the model to have the option to be ordered in Semi-Automatic and Fully Automatic as required? A unit cannot toggle between semi-automatic and fully-automatic, this required to be chosen at the time of order.

Answer: The AED can be Semi-Automatic or Fully Automatic. We do not expect the AED to have both capabilities

3. Mandatory Requirement: "The AED must have a metronome and CPR coaching. CPR coaching is described as the ability to provide instructions for Adult and Pediatric CPR, including feedback when no CPR is detected, rate guidance, and a metronome."
We ask that this requirement be changed to make the unit mandatory to provide "CPR Feedback that provides depth and rate feedback" to ensure proper CPR is being given with prompts telling the user to push harder when required. When referring to AEDs, the term CPR Coaching generally refers to instruction of how to do CPR rather than CPR Feedback.

Answer: We will accept AEDs with CPR Feedback.

4. Table 1 of 2: Initial Requirement "Item 2 and 3"
Can you please clarify that item 2 and 3 are extra and in addition to the batteries and pads included with the AED kits of Item 1, and the batteries that would already be included in the Cabinet of item 4. AEDs cannot be sold without pads and batteries, and would like to clarify that it is asking for an additional 80 sets of pads and battery combos.

Answer: Items 2 is extra to the battery and pads that are included with the AED. Item 3 is the cost for batteries for the cabinet if not included with the cabinet.

Please note the following change to Statement of Requirement :

Environment Specifications

The AED must be capable of a storage temperature ranging of 0°C to 30°C, with electrodes and battery.

ALL OTHER TERMS AND CONDITIONS REMAIN UNCHANGED