

**RETURN BIDS TO:
RETOURNER LES SOUMISSIONS À:**

Bid Receiving - PWGSC
11 Laurier St.
Place du Portage , Phase III
Core 0B2
Gatineau, Québec K1A 0S5
Bid Fax: (819) 997-9776
ePost Connect Email:
tpsgc.dgareceptiondessoumissions-abbidReceiving.pwgsc@tpsgc-pwgsc.gc.ca

**SOLICITATION AMENDMENT
MODIFICATION DE L'INVITATION**

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Solicitation remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

**Vendor/Firm Name and Address
Raison sociale et adresse du
fournisseur/de l'entrepreneur**

Issuing Office - Bureau de distribution

Innovation Procurement Directorate
Direction des achats innovateurs
Terrasses de la Chaudière
10, rue Wellington, Étage 4
Gatineau, Québec
K1A 0S5

Title-Sujet Area Detection and Identification System (ADIS)	
Solicitation No. - N° de l'invitation W8476-18ADIS/B	Amendment No. - N° modif. 020
Client Reference No. - N° de référence du client W8476-18ADIS/B	Date 3 June 2020
GETS Reference No. - N° de référence de SEAG	
File No. - N° de dossier 010sl.W8476-18ADIS/B	CCC No./N° CC – FMS NO. / N° VME
Solicitation Closes – L'invitation prend fin at – à 2:00 PM on – le 30 July 2020	Time Zone Fuseau horaire Eastern Standard Time EST
F.O.B. – F.A.B	
Plant-Usine : <input type="checkbox"/> Destination: <input type="checkbox"/> Other-Autre: <input type="checkbox"/>	
Address Enquiries to: - Adresser toutes questions à: April Campbell	Buyer Id – Id de l'acheteur 010SL
Telephone No. - N° de téléphone 613-858-9485	FAX No. - N° de FAX
Destination of Goods, Services and Construction: Destinations des biens, services et construction :	
Specified Herein Précisé dans les présentes	

Instructions : See Herein

Instructions : voir aux présentes

Delivery Required - Livraison exigée See Herein	Delivery Offered - Livraison proposée
Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur	
Telephone No. - N° de telephone Facsimile No. - N° de télécopieur	
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print)	
Nom et titre de la personne autorisée à signer au nom du fournisseur/de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date

THIS AMENDMENT IS RAISED TO RESPOND TO BIDDER QUESTIONS

- Q135 In Amendment 017, question Q110 and A110, claims that Canada can provide the 8 spectra of the test items. In Amendment 005 it is stated “If a bidder does not have the required spectra, Canada can help in providing the signatures at different spectral resolutions, and also in different data format (ASCII text files, binary, etc.)” How does a potential contractor go about requesting and receiving the 8 spectra of the test items? Who should we contact? What is the preferred method of contact (ie. email, phone, etc.)?
- A135 Canada has provided the spectra zip file as an attachment called ADIS_test_spectra.zip on Buy and Sell.
- Q136 Table A in Appendix AA1 to Annex A: As proposed in the answer provided in Amendment 5, Answer 14 (b), can the PWGSC provide IR Spectra for all required agents?
- A136 DND will be using 8 simulants, the spectra for which can be found separately on the Buy and Sell site. The ReadMe documents in the zip folders provide additional information about the spectra files. Please refer to A135 above.
- Q137 Re Vol 2, Appendix AA: It is clear that the battery pack and charger must meet the requirements under section 3.2.4 Power. Please confirm the battery pack and charger do not need to meet all of the requirements of section 3.4 Specialty Engineering and 3.5 Environmental Characteristics ?
- A137 The battery pack and battery charger are considered part of the ADIS unit and will need to meet the specifications in sections 3.4 and 3.5.
- Q138 Attachment 2A Part 4 DRDC Lab testing states that 5 test substances, from the list of 8 possible test substances, will be used for the testing, with the possibility of a Nitrogen filled cylinder being included as one of the five. Please confirm that all bidders systems will be tested using the same 5 test substances to ensure fairness? Please also confirm that the five test substances will be chosen prior to bid closing.
- A138 Canada will use the same 4 test substance and 1 blank cylinders to evaluate every competing Bidder system. Bidders must be prepared to have their systems detect all 8 substances as Bidders will not be informed beforehand which substances will be part of the trial. The substances will be chosen prior to the start of testing but after bid closing.