



RETURN BIDS TO:

RETOURNER LES SOUMISSIONS À:

Travaux publics et Services gouvernementaux
Canada

Voir dans le document/

See herein

NA

Québec

NA

**SOLICITATION AMENDMENT
MODIFICATION DE L'INVITATION**

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Solicitation remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

Vendor/Firm Name and Address

Raison sociale et adresse du
fournisseur/de l'entrepreneur

Issuing Office - Bureau de distribution

Travaux publics et Services gouvernementaux Canada
Place Bonaventure, portail Sud-Oue
800, rue de La Gauchetière Ouest
7^e étage, suite 7300
Montréal
Québec
H5A 1L6

Title - Sujet Field Flow Fractionation system	
Solicitation No. - N° de l'invitation K8D10-210111/A	Amendment No. - N° modif. 001
Client Reference No. - N° de référence du client K8D10-210111	Date 2020-10-07
GETS Reference No. - N° de référence de SEAG PW-SMTA-070-15863	
File No. - N° de dossier MTA-0-43088 (070)	CCC No./N° CCC - FMS No./N° VME
Solicitation Closes - L'invitation prend fin at - à 02:00 PM on - le 2020-11-09	Time Zone Fuseau horaire Heure Normale du l'Est HNE
F.O.B. - F.A.B. Plant-Usine: <input type="checkbox"/> Destination: <input type="checkbox"/> Other-Autre: <input type="checkbox"/>	
Address Enquiries to: - Adresser toutes questions à: Therien, Renée-Ann	Buyer Id - Id de l'acheteur mta070
Telephone No. - N° de téléphone (514) 703-4947 ()	FAX No. - N° de FAX () -
Destination - of Goods, Services, and Construction: Destination - des biens, services et construction:	

Instructions: See Herein

Instructions: Voir aux présentes

Delivery Required - Livraison exigée	Delivery Offered - Livraison proposée
Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur	
Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur	
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date

AMENDMENT 001:

This amendment is raised to update the following items of the solicitation document:

- Item 2.8 of annex A and annex C of the solicitation document;
- Item 3.2 of annex A of the solicitation document;
- Items 6.1.1 and 6.1.2 of annex A of the solicitation document;

DELETE:

Item 2.8 of annex A and annex C:

2.8 The fractionation unit must be able to operate with a Dionex brand pump (ICS-3000) and a Dionex brand sampler (ICS-3000).

Item 3.2 of annex A:

3.2 The software must be compatible with a Dionex pump (ICS-3000) and a Dionex brand sampler (ICS-3000).

Items 6.1.1 and 6.1.2 of annex A:

6.1.1 The flow fractionation system must be able to operate by means of a single isocratic pump for all flows (inlet, transverse and channel).

6.1.2 The adjustable pump flow must be able to vary from 0.001 to 10 mL / min. at pressures up to 600 bar. The pump must be in metal free version. Such pump could be optional.

INSERT:

Item 2.8 of annex A and annex C:

2.8 The laboratory already has a metal-free pump and a metal-free sampler. The pump that we have allows a flow varying from 0.001 to 10.0 mL / min and can reach a maximum pressure of 35 MPa. The sampler we have includes a loop injection system and can inject volumes ranging from 1 to 100µL.

If the bidder offers us a flow-through fractionation unit that works with a single pump and the

Solicitation No. - N° de l'invitation
K8D10-210111/A
Client Ref. No. - N° de réf. du client
K8D10-210111

Amd. No. - N° de la modif.
001
File No. - N° du dossier
MTA-0-43088

Buyer ID - Id de l'acheteur
MTA070
CCC No./N° CCC - FMS No./N° VME

sampler we own, the latter must not include a pump and a sampler in their bid. The supplier must indicate to us in his quote with which pump and which sampler the proposed unit is compatible with.

If the fractionation unit cannot operate with only one pump, the tenderer can offer us a fractionation unit with its sampler, degasser and its 2 pump (s) necessary for its use.

Item 3.2 of annex A:

3.2 The operating software must allow the control of the components (pumps, sampler) of the proposed fractionation system.

Item 6.1.1 of annex A:

6.1 Isocratic pump

The adjustable flow rate of the pump must be able to vary from 0.001 to 10 mL / min. at pressures up to 400 bar. The pump must come with the metal free option. Such a pump is requested as an option.

ALL OTHER TERMS AND CONDITIONS REMAIN THE SAME.