

	National Defence Défense Nationale		Back to the DID List
DATA ITEM DESCRIPTION - DESCRIPTION DE DONNÉES			
1. TITLE – TITRE		2. IDENTIFICATION NUMBER - NUMÉRO D'IDENTIFICATION	
TEST EQUIPMENT MAINTENANCE PLAN		DID 1.208.1	
3. DESCRIPTION / PURPOSE – DESCRIPTION / OBJET			
The Test Equipment Maintenance Plan shall provide a detailed description of the organization, facilities and procedures to be implemented by the Contractor to ensure that all test equipment is repaired as required and is calibrated at the required intervals and that all sites have calibrated test equipment available at all times.			
4. APPROVAL DATE DATE D'APPROBATION	5. OFFICE OF PRIMARY INTEREST (OPI) BUREAU DE PREMIERE RESPONSABILITÉ (BPR)	6. GIDEP APPLICABLE D'ÉCHANGE DE DONNÉES PERTINENT	
August 2020	NWSO Technical Authority (TA)	N/A	
7. APPLICATION / INTERRELATIONSHIP – APPLICATION / INTERDÉPENDANCE			
CDRL-1.208.1 and SOW paragraphs 1.208.1, 3.23, 3.25 refer. This DID contains the format and content preparation instructions for the data generated under the work tasks described in the NWS O&M SOW.			
8. ORIGINATOR - AUTEUR		9. APPLICABLE FORMS - FORMULES PERTINENTES	
NWSO TA		NIL	
10. PREPARATION INSTRUCTIONS – INSTRUCTIONS SUR LA PRÉSENTATION DES DONNÉES			
<div> <div>10.1</div> <div> <u>Source Document</u> NWS O&M SOW Section 3, paragraphs 1.208.1, Section 3, paragraphs 3.23, 3.25. </div> </div> <div> <div>10.2</div> <div> <u>Content and Format</u> <div>10.2.1</div> <div> The Test Equipment Maintenance Plan shall be prepared and delivered in Contractor format. The Plan shall describe in detail the Contractor's method of calibrating or having calibrated, and of repairing or having repaired, all test equipment. The Plan shall include: <ul style="list-style-type: none"> a. Responsibilities of Contractor personnel for test equipment calibration at all NWS facilities; b. Method of authorizing test equipment/instruments for all sites and facilities; c. Facilities to be used for repair of test equipment; d. Facilities to be used for various levels of calibration; e. Procedures for transporting test equipment to/from calibration or repair facilities; f. Procedures for rotating spare test equipment to ensure that each site has full complement of authorized test equipment at all times; and g. Documentation such as labels, forms, and messages (with instructions for completion of each) to control and record the location and status of all test equipment. </div> </div> <div>10.2.2</div> <div> The Test Equipment Maintenance Plan shall include the following four (4) reports: <ul style="list-style-type: none"> a. Test Equipment Master List; b. Test Equipment Calibration Report; c. Test Equipment Calibration Non-Compliance Report; and d. Test Equipment Repair Report. </div> </div>			

10.2.3 Test Equipment Master List

The List shall be prepared and delivered quarterly in Contractor format. The Contractor shall provide a Master List of all Test Equipment for which the Contractor is responsible. The List shall include the following:

- a. Site;
- b. NSN;
- c. Nomenclature;
- d. Serial number;
- e. Calibration interval;
- f. Date of last calibration; and
- g. Date next calibration due.

Test Equipment Master List shall be organized by NSN and by Site/Location.

10.2.4 Test Equipment Calibration Report

The Report shall be prepared and delivered quarterly in Contractor format. The Report shall document calibration of test equipment for which the Contractor is responsible. The Report shall include the following:

- a. Equipment nomenclature;
- b. Equipment model;
- c. Equipment serial number;
- d. Date of calibration;
- e. Calibration facility;
- f. Date next calibration due; and
- g. Location (after calibration).

10.2.5 Test Equipment Calibration Non-Compliance Report

The Report shall be prepared and delivered quarterly in Contractor format. The Report shall provide a summary of test equipment items not calibrated in accordance with the schedule of requirements of the Test Equipment Maintenance Plan. The Report shall include:

- a. Equipment nomenclature;
- b. Equipment model;
- c. Equipment serial number;
- d. Date calibration due;
- e. Calibration facility;
- f. Date calibration rescheduled; and
- g. Reason for non-compliance.

10.2.6 Test Equipment Repair Report

The Report shall be prepared and delivered quarterly in Contractor format. The Report is a summary of repairs to NWS test equipment. The Report shall include:

- a. Equipment nomenclature;
- b. Equipment model;
- c. Equipment serial number;
- d. Date sent for repair;
- e. Repair facility; and
- f. Date received from repair facility.