

		National Defence Défense Nationale	<a href="#">Back to the DID List</a>
<b>DATA ITEM DESCRIPTION - DESCRIPTION DE DONNÉES</b>			
<b>1. TITLE – TITRE</b>		<b>2. IDENTIFICATION NUMBER - NUMÉRO D'IDENTIFICATION</b>	
Test Equipment Calibration Non-Compliance Report		DID 4.44	
<b>3. DESCRIPTION / PURPOSE – DESCRIPTION / OBJET</b>			
The Report shall be prepared and made available on-line quarterly in Contractor format. The Report shall provide a summary of test equipment items not calibrated in accordance with the schedule of requirements.			
<b>4. APPROVAL DATE DATE D'APPROBATION</b>	<b>5. OFFICE OF PRIMARY INTEREST (OPI) BUREAU DE PREMIERE RESPONSABILITÉ (BPR)</b>	<b>6. GIDEP APPLICABLE D'ÉCHANGE DE DONNÉES PERTINENT</b>	
September 2020	NWSO Technical Authority (TA)		
<b>7. APPLICATION / INTERRELATIONSHIP – APPLICATION / INTERDÉPENDANCE</b>			
CDRL 4.44 and SOW paragraph 4.44 refer. This DID contains the format and content preparation instructions for the data generated under the work tasks described in the NWS O&M SOW.			
<b>8. ORIGINATOR - AUTEUR</b>		<b>9. APPLICABLE FORMS - FORMULES PERTINENTES</b>	
NWSO TA			
<b>10. PREPARATION INSTRUCTIONS – INSTRUCTIONS SUR LA PRÉSENTATION DES DONNÉES</b>			
10.1 <u>Source Document</u> NWS O&M SOW Section 4, paragraph 4.44.  10.2 <u>Content and Format</u> 10.2.2 The Calibration Non-Compliance Report shall include the following: 10.2.2.1 Equipment nomenclature; 10.2.2.2 Equipment model; 10.2.2.3 Equipment serial number; 10.2.2.4 Date of calibration; 10.2.2.5 Calibration facility; 10.2.2.6 Date calibration due rescheduled; and 10.2.2.7 Reason for non-compliance.			
<div style="text-align: right;"> Page 1 of 1  September 2009 </div>			