



**RETURN BIDS TO:**

**RETOURNER LES SOUMISSIONS À:**

Bid Receiving Public Works and Government  
Services Canada/Réception des soumissions/Travaux  
publics et Services gouvernementaux Canada  
See herein for bid submission  
instructions/

Voir la présente pour les  
instructions sur la présentation  
d'une soumission

NA  
Ontario

**SOLICITATION AMENDMENT  
MODIFICATION DE L'INVITATION**

The referenced document is hereby revised; unless otherwise  
indicated, all other terms and conditions of the Solicitation  
remain the same.

Ce document est par la présente révisé; sauf indication contraire,  
les modalités de l'invitation demeurent les mêmes.

**Comments - Commentaires**

**Vendor/Firm Name and Address**  
**Raison sociale et adresse du**  
**fournisseur/de l'entrepreneur**

**Issuing Office - Bureau de distribution**  
Public Works and Government Services / Travaux  
publics et services gouvernementaux  
Kingston Procurement  
Des Acquisitions Kingston  
86 Clarence Street, 2nd floor  
Kingston  
Ontario  
K7L 1X3

<b>Title - Sujet</b> EMI Chamber Equipment	
<b>Solicitation No. - N° de l'invitation</b> W3474-211561/A	<b>Amendment No. - N° modif.</b> 003
<b>Client Reference No. - N° de référence du client</b> W3474-21-1561	<b>Date</b> 2020-12-17
<b>GETS Reference No. - N° de référence de SEAG</b> PW-\$KIN-610-8193	
<b>File No. - N° de dossier</b> KIN-0-54049 (610)	<b>CCC No./N° CCC - FMS No./N° VME</b>
<b>Solicitation Closes - L'invitation prend fin</b> <b>at - à 02:00 PM</b> Eastern Standard Time EST <b>on - le 2020-12-21</b> Heure Normale de l'Est HNE	
<b>F.O.B. - F.A.B.</b>	
<b>Plant-Usine:</b> <input type="checkbox"/> <b>Destination:</b> <input checked="" type="checkbox"/> <b>Other-Autre:</b> <input type="checkbox"/>	
<b>Address Enquiries to: - Adresser toutes questions à:</b> Correia-Reid, Vincent	<b>Buyer Id - Id de l'acheteur</b> kin610
<b>Telephone No. - N° de téléphone</b> (873) 355-6115 ( )	<b>FAX No. - N° de FAX</b> (613) 545-8067
<b>Destination - of Goods, Services, and Construction:</b> <b>Destination - des biens, services et construction:</b> Aerospace and Telecommunication Engineering Support Squadron 8 Wing/CFB Trenton Astra, ON K0K 3W0	

**Instructions: See Herein**

**Instructions: Voir aux présentes**

<b>Delivery Required - Livraison exigée</b>	<b>Delivery Offered - Livraison proposée</b>
<b>Vendor/Firm Name and Address</b> <b>Raison sociale et adresse du fournisseur/de l'entrepreneur</b>	
<b>Telephone No. - N° de téléphone</b> <b>Facsimile No. - N° de télécopieur</b>	
<b>Name and title of person authorized to sign on behalf of Vendor/Firm</b> <b>(type or print)</b> <b>Nom et titre de la personne autorisée à signer au nom du fournisseur/</b> <b>de l'entrepreneur (taper ou écrire en caractères d'imprimerie)</b>	
<b>Signature</b>	<b>Date</b>

---

### AMENDMENT 003

This Amendment 003 is raised to extend the closing date of the RFP to December 21, 2020 and respond to inquiries.

Q1. Is it possible to submit a bid without a price for every item in Annex "B"?

A1. Per the Basis of Selection under 4.2.1 of the RFP: "The responsive bid with the lowest evaluated price on an item by item basis will be recommended for award of a contract." Therefore, a bid can be accepted even if it does not contain pricing for every item.

Q2. Due to furloughs associated with Covid-19 several component manufacturers are at a reduced work force and they cannot deliver the parts until March 31, 2021. It is impossible to deliver some of the requested products until end of March 2021. We kindly request to remove mandatory delivery time of March 31, 2021 and replace it with: "it is desired to receive all the deliverables before March 31, 2021 but delivery after this date is also acceptable"

A2. The delivery deadline is a client mandatory requirement and will not be modified at this time.

Q3. In Annex "B" Some items of the table (for example 3.4 & 3.5) have quantity of more than one and are consisting of different products while there is a column in the table titled "unit price". It seems that total extended price is required for this column. We kindly request to replace "unit price" with "total extended price".

A3. The prices requested in Annex "B" cover the entire applicable section in the Requirement, and it may be the case that the all-inclusive "unit price" covers more than one item. However, the Basis of Selection specifies "The evaluated prices are the unit prices provided for each item.", therefore the wording in Annex "B" will not be changed.

Q4. Where can I put the freight cost on the bid?

A4. Unit prices must be placed in the Basis of Payment for the applicable item, and the unit prices must be all-inclusive, excluding applicable taxes.

Q5. After reviewing the RFP, we understand that the criteria for choosing each listed item is (in order of importance): 1) Technology, and 2) Price. Based on this, what are the deciding factors if the item does not meet all of the listed technical criteria?

A5. There is no technical evaluation as part of the award process, only price is considered, but all goods provided must meet the Requirement at Annex A or they will be rejected upon delivery.

Q6. We were wondering if this system is currently using any versions of TDK software. If so, could you let us know what version you are using?

A6. We are using TDK Radiated Immunity Lab (RAD-LAB) version 10.93.0.1.

Q7. Your Technical Specifications and Requirements (Section 3, page 12 of 30) does not list any power meters as part of the equipment requirements. Typically CS-114/RS-103 testing would utilize power meters to monitor the power output of the amplifiers. Will you need power meters for this application?

A7. Section 3.2.5.3 Forward and Reflected Power Display states that all amplifiers must:

Display forward and reflected power with at least 3 significant digits.

Solicitation No. - N° de l'invitation

W3474-211561/A

Client Ref. No. - N° de réf. du client

W3474-211561

Amd. No. - N° de la modif.

003

File No. - N° du dossier

KIN-0-54049

Buyer ID - Id de l'acheteur

kin610

CCC No./N° CCC - FMS No./N° VME

---

No additional power meter is necessary. It is the electric field that is monitor externally and adjusted by the software to set values.

**ALL OTHER TERMS AND CONDITIONS OF THE RFP REMAIN UNCHANGED**