



**RETURN OFFERS TO:
RETOURNER LES OFFRES A:**

Attention: Candice Therien
Email: Candice.therien@rcmp-grc.gc.ca

Note - RCMP has the following e-mail restrictions:
The maximum e-mail message size is 5 MB.
Zip files not accepted.

**AMENDMENT TO THE
REQUEST FOR A STANDING
OFFER**

**MODIFICATION À LA
DEMANDE D'OFFRE À
COMMANDES**

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Request for a Standing Offer remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de la demande d'offre à commandes demeurent les mêmes.

Comments: - Commentaries :

Title – Sujet Blood Alcohol Collection Kits		Date December 17th, 2020
Solicitation No. – N° de l'invitation 202102756A		Amendment No. – N° de la modification 001
Client Reference No. - No. De Référence du Client 202102756		
Solicitation Closes – L'invitation prend fin		
At / à :	14 :00	EST(Eastern StandardTime) HAE (heure avancée de l'Est)
On / le :	January 5th, 2020	
Delivery - Livraison See herein — Voir aux présentes	Taxes - Taxes See herein — Voir aux présentes	Duty – Droits See herein — Voir aux présentes
Destination of Goods and Services – Destinations des biens et services See herein — Voir aux présentes		
Instructions See herein — Voir aux présentes		
Address Inquiries to – Adresser toute demande de renseignements à Candice.therien@rcmp-grc.gc.ca		
Telephone No. – No. de téléphone		Facsimile No. – No. de télécopieur
Delivery Required – Livraison exigée See herein — Voir aux présentes		Delivery Offered – Livraison proposée
Vendor/Firm Name, Address and Representative – Raison sociale, adresse et représentant du fournisseur/de l'entrepreneur:		
Telephone No. – No. de telephone		Facsimile No. – No. de télécopieur
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) – Nom et titre de la personne autorisée à signer au nom du fournisseur/de l'entrepreneur (taper ou écrire en caractères d'imprimerie)		
Signature		Date

This amendment is raised to address the following:

- To respond to questions received during the RFSO period

QUESTIONS AND ANSWERS

Q1) Section 2.2 & 2.3 require a blood tube holder and needle, with a safety shield, that is 21 G x 1.5". Becton Dickinson produces these items but no longer make the requested size but have a 21 G x 1.25" model. Would this be an acceptable replacement?

A1) After consultation, we have confirmed that the specified replacement would be acceptable and should not present any issues.