



RETURN BIDS TO:

RETOURNER LES SOUMISSIONS À:

Travaux publics et Services gouvernementaux
Canada

Voir dans le document/

See herein

NA

Québec

NA

**SOLICITATION AMENDMENT
MODIFICATION DE L'INVITATION**

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Solicitation remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

Vendor/Firm Name and Address

Raison sociale et adresse du
fournisseur/de l'entrepreneur

Issuing Office - Bureau de distribution

Travaux publics et Services gouvernementaux Canada
Place Bonaventure, portail Sud-Oue
800, rue de La Gauchetière Ouest
7^e étage, suite 7300
Montréal
Québec
H5A 1L6

Title - Sujet Wireless Monitoring System Wireless Monitoring System (temperature & CO2 concentration)	
Solicitation No. - N° de l'invitation 39903-210584/A	Amendment No. - N° modif. 003
Client Reference No. - N° de référence du client 39903-210584	Date 2021-01-06
GETS Reference No. - N° de référence de SEAG PW-SMTA-030-15973	
File No. - N° de dossier MTA-0-43204 (030)	CCC No./N° CCC - FMS No./N° VME
Solicitation Closes - L'invitation prend fin at - à 02:00 PM Eastern Standard Time EST on - le 2021-01-21 Heure Normale du l'Est HNE	
F.O.B. - F.A.B. Specified Herein - Précisé dans les présentes Plant-Usine: <input type="checkbox"/> Destination: <input type="checkbox"/> Other-Autre: <input checked="" type="checkbox"/>	
Address Enquiries to: - Adresser toutes questions à: Pommet, Bruno André	Buyer Id - Id de l'acheteur mta030
Telephone No. - N° de téléphone (514) 702-9582 ()	FAX No. - N° de FAX () -
Destination - of Goods, Services, and Construction: Destination - des biens, services et construction:	

Instructions: See Herein

Instructions: Voir aux présentes

Delivery Required - Livraison exigée	Delivery Offered - Livraison proposée
Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur	
Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur	
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date

AMENDMENT 003

This amendment is to publish a new series of Questions and Answers (Q&A) related to the Requirement.

Question 5: In reference with amendment 002 (Grid – Annex A2): when there are two numbers in red, does this mean two sensors are required because there are two chambers?

Answer 5: No. This means that the probe associated with this equipment must be calibrated at these two temperatures in red.

Question 6: In reference with amendment 002 (Grid – Annex A2), it says 'Calibration required' in the 2nd column. Does it mean the sensors must be calibrated in the equipment and not be pre-calibrated with a certificate?

Answer 6: No. This means the probe associated with this equipment must be pre-calibrated with a certificate. The probe must be calibrated by an ISO / IEC 17025 Standard accredited calibration provider.

Question 7: Ref Annex A - criteria 4.17 & 4.18

Question 7.1: Is the entire temperature range (-150 ° C and 65 ° C) to be covered by each and every individual probe?

Answer 7.1: No. The individual probes must only cover the temperatures of the associated equipment (see Annex A2 – Amendment 002). It is the set of all temperatures of all the equipment to be monitored which is in the range of -150 ° C to 65 ° C. Note: although no temperature of 65 ° C is indicated in the equipment list of Annex A2, the temperature range indicated in criterion 4.17 of the Annex A goes up to 65 ° C, in order to meet an occasional and future need.

Question 7.2: Can you clarify the purpose of Annex A2, posted as part of the amendment 002? Can you confirm the following?

- The probes are no longer required to cover the entire range (-150 ° C and 65 ° C). Instead, they need to comply with the values shown on Annex 2.
- The temperature accuracy requirement was amended from 0.1° C to 0.5° C.
- The CO2 accuracy requirement was amended from 0.1% to 1%.

Answer 7.2: The purpose of Annex A2 is to provide a detailed equipment list to be monitored with the exact quantity, type of equipment, the monitored measurement with its acceptable range, and its location (room number).

- The individual probes do not have to cover the entire range (-150°C and 65°C). The individual probes must only cover the temperatures of the associated equipment (see Annex A2). It is the set of all temperatures of all the equipment to be monitored which is in the range

of -150°C to 65°C. Note: although no temperature of 65 ° C is indicated in the equipment list of Annex A2, the temperature range indicated in 4.17 of the Annex A goes up to 65 ° C, in order to meet an occasional and future need.

- No. The temperature accuracy requirement of the probe is still 0.1° C. The range indicated in the list of equipment (see Annex A2) only shown the range of temperature or % CO2 allowed for each monitored measurement. The range is related to the monitored measurement, not to the precision of the probe.
- No. The CO2 accuracy requirement of the probe is still 0.1° C. The range indicated in the list of equipment (see Annex A2) only shown the range of temperature or % CO2 allowed for each monitored measurement. The range is related to the monitored measurement, not to the precision of the probe.

ALL OTHER TERMS AND CONDITIONS REMAIN THE SAME.