



RETURN BIDS TO:

RETOURNER LES SOUMISSIONS À:

Travaux publics et Services gouvernementaux
Canada
Voir dans le document/
See herein
NA
Québec
NA

**SOLICITATION AMENDMENT
MODIFICATION DE L'INVITATION**

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Solicitation remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

**Vendor/Firm Name and Address
Raison sociale et adresse du
fournisseur/de l'entrepreneur**

Issuing Office - Bureau de distribution
Travaux publics et Services gouvernementaux Canada
Place Bonaventure, portail Sud-Oue
800, rue de La Gauchetière Ouest
7e étage, suite 7300
Montréal
Québec
H5A 1L6

Title - Sujet Wireless Monitoring System Wireless Monitoring System (temperature & CO2 concentration)	
Solicitation No. - N° de l'invitation 39903-210584/A	Amendment No. - N° modif. 004
Client Reference No. - N° de référence du client 39903-210584	Date 2021-01-13
GETS Reference No. - N° de référence de SEAG PW-SMTA-030-15973	
File No. - N° de dossier MTA-0-43204 (030)	CCC No./N° CCC - FMS No./N° VME
Solicitation Closes - L'invitation prend fin at - à 02:00 PM Eastern Standard Time EST on - le 2021-01-21 Heure Normale de l'Est HNE	
F.O.B. - F.A.B. Specified Herein - Précisé dans les présentes	
Plant-Usine: <input type="checkbox"/> Destination: <input type="checkbox"/> Other-Autre: <input checked="" type="checkbox"/>	
Address Enquiries to: - Adresser toutes questions à: Pomet, Bruno André	Buyer Id - Id de l'acheteur mta030
Telephone No. - N° de téléphone (514) 702-9582 ()	FAX No. - N° de FAX () -
Destination - of Goods, Services, and Construction: Destination - des biens, services et construction:	

Instructions: See Herein

Instructions: Voir aux présentes

Delivery Required - Livraison exigée	Delivery Offered - Livraison proposée
Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur	
Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur	
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date

AMENDMENT 004

This amendment is to publish a new series of Questions and Answers (Q&A) related to the Requirement and to modify the Requirement (Annex A).

Question 8: For a server-based monitoring software the users will be able to access the software via an internet connection if the Canadian Food Inspection Agency (CFIA) provides them with VPN login. Will VPN access be provided?

Answer 8: Yes. VPN access is provided to all CFIA users by the CFIA.

Question 9: If Cloud software is acceptable, Users can access the software directly with an internet connection using a secure login. Is Cloud storage and software interface an acceptable alternative to in-house server?

Answer 9: Please refer to Solicitation document Amendment 002, question/answer #3.

Question 10: Is a NIST traceable calibration certificate acceptable or does the calibration certificate have to be ISO 17025?

Answer 10: Please refer to Solicitation document (000), Annex A, criterion 4.21. It is clearly indicated that probes must be calibrated by a minimum of one accredited calibration provider to meet the requirements of ISO / IEC 17025.

Question 11: Must or can a written proposal be included with the quotation?

Answer 11: Yes. Please refer to Solicitation document (000) Part 3 – Bid Preparation Instructions.

Question 12: What is the installation timeframe? Start date? End date? Will work be expected to be completed on weekends? After hours? Or during regular business hours?

Answer 12: Please refer to Solicitation document (000) Part 2, section 2.1.2. Delivery/installation is requested as soon as possible. All installation details will be given once the contract is awarded to the winning bidder.

Question 13: Is local support in the Montreal area important or required? Do you require English/French service personnel?

Answer 13: Please refer to the new section 8 of Annex A (see below).

Question 14: Do you require 24/7 response time? If not, what is the expected response time?

Solicitation No. - N° de l'invitation
39903-210584/A
Client Ref. No. - N° de réf. du client
39903-210584

Amd. No. - N° de la modif.
004
File No. - N° du dossier
MTA-0-43204

Buyer ID - Id de l'acheteur
MTA030
CCC No./N° CCC - FMS No./N° VME

Answer 14: Please refer to the new section 8 of Annex A (see below).

Question 15: Do alarm notification also need to be generated by CALL SMS/TEXT?

Answer 15: No. This is not required.

Question 16: Do you want to monitor Webcams?

Answer 16: No.

Question 17: Do you want to monitor any 3rd party sensors? Gas detection? Ph?

Answer 17: No, as indicated in the Annex A, the required system has to monitor and record the temperatures and CO2 concentrations of laboratory equipment and premises of the CFIA.

Question 18: Does the system need to be validated? IQ/OQ/PQ?

Answer 18: No

Question 19: Does the software need be FDA CFR 21 Part 11? GAMP5?

Answer 19: No

Question 20: Will this project be reusing any existing sensors?

Answer 20: No.

Question 21: Do you require color-coded alarm bands to be visible on the chart view of the data for quick analysis of alarm conditions?

Answer 21: No. This is not required.

Question 22: Are Travel & Living expenses to be included for startup and commissioning in the quote?

Answer 22: Please refer to Solicitation document (000), Annex B (Basis of payment), line items 4 and 5.

**DELETE:
ANNEX A**

4. Mandatory technical specifications

4.18 Temperature probes must have an accuracy and resolution less than or equal to 0.1°C and CO2 probes must have an accuracy and resolution less than or equal to 0.1%.

**INSERT:
ANNEX A**

4. Mandatory technical specifications

4.18 Temperature probes must have an accuracy less than or equal to 1°C and a resolution less than or equal to 0.1°C. CO2 probes must have an accuracy less than or equal to 1°C and a resolution less than or equal to 0.1%.

**INSERT:
ANNEX A**

8. Technical Assistance Service

8.1 The supplier must provide a technical assistance service upon the installation and commissioning of the system, in French, by telephone and/or email during the regular business hours and from Monday to Friday.

8.2 A 24-hour response time is required (phone and/or email).

DELETE:

ANNEX C - MANDATORY TECHNICAL CRITERIA TO BE DEMONSTRATED

4.18	Temperature probes must have an accuracy and resolution less than or equal to 0.1°C and CO2 probes must have an accuracy and resolution less than or equal to 0.1%.	
------	---	--

INSERT:

ANNEX C - MANDATORY TECHNICAL CRITERIA TO BE DEMONSTRATED

4.18	Temperature probes must have an accuracy less than or equal to 1°C and a resolution less than or equal to 0.1°C. CO2 probes must have an accuracy less than or equal to 1°C and a resolution less than or equal to 0.1%.	
------	--	--

ALL OTHER TERMS AND CONDITIONS REMAIN THE SAME.