



RETURN BIDS TO:

RETOURNER LES SOUMISSIONS À:

Bid Receiving Public Works and Government
Services Canada/Réception des soumissions/Travaux
publics et Services gouvernementaux Canada
See herein for bid submission
instructions/

Voir la présente pour les
instructions sur la présentation
d'une soumission

NA
Ontario

SOLICITATION AMENDMENT MODIFICATION DE L'INVITATION

The referenced document is hereby revised; unless otherwise
indicated, all other terms and conditions of the Solicitation
remain the same.

Ce document est par la présente révisé; sauf indication contraire,
les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

Vendor/Firm Name and Address
Raison sociale et adresse du
fournisseur/de l'entrepreneur

Issuing Office - Bureau de distribution
Public Works and Government Services / Travaux
publics et services gouvernementaux
Kingston Procurement
Des Acquisitions Kingston
86 Clarence Street, 2nd floor
Kingston
Ontario
K7L 1X3

Title - Sujet Wearable Eye Tracking Suite	
Solicitation No. - N° de l'invitation W7719-215523/A	Amendment No. - N° modif. 002
Client Reference No. - N° de référence du client W7719-21-5523	Date 2021-03-18
GETS Reference No. - N° de référence de SEAG PW-\$KIN-930-8355	
File No. - N° de dossier KIN-0-54170 (930)	CCC No./N° CCC - FMS No./N° VME
Solicitation Closes - L'invitation prend fin at - à 02:00 PM Eastern Daylight Saving Time EDT on - le 2021-03-31 Heure Avancée de l'Est HAE	
F.O.B. - F.A.B. Specified Herein - Précisé dans les présentes Plant-Usine: <input type="checkbox"/> Destination: <input type="checkbox"/> Other-Autre: <input checked="" type="checkbox"/>	
Address Enquiries to: - Adresser toutes questions à: Littlefield, Mike	Buyer Id - Id de l'acheteur kin930
Telephone No. - N° de téléphone (613) 449-4206 ()	FAX No. - N° de FAX (613) 545-8067
Destination - of Goods, Services, and Construction: Destination - des biens, services et construction: Defence R&D Canada - Toronto 1133 Sheppard Avenue West Toronto, ON M3K 2C9	

Instructions: See Herein

Instructions: Voir aux présentes

Delivery Required - Livraison exigée	Delivery Offered - Livraison proposée
Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur	
Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur	
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date

Amendment 002 – is being issued in response to Offeror's questions and Canada's responses.

Q1: Our company offers both a 50 and 100 Hz system, so either system meets the requirement. Does the end user want the 100 Hz, or pricing for both the 50 and 100 Hz?

A1: It's the supplier's decision, the requirement indicates 50hz is the acceptable minimum and more points are awarded should the Hz increase.

Q2: Analysis software specs – "Blink duration and frequency data" – Our company's software provides the pupil data for these metrics but it is not an automatic output like other automated pupil data output. It requires an algorithm (such as PyTrack-NTU) in post-processing.

A2: It is acceptable for the "Blink duration and frequency data" feature of the analysis software to be available via additional processing.

Q3: Pricing Basis A and B – A1 and B1 and B1 and B2 – our company's glasses include the Recording unit. A1/B1 will include 12 Recording devices. Is A2/B2 requesting 18 additional recording units or 6? If the latter (6), can the A2/B2 line item request be amended to Qty 6 for pricing clarification?

A3: Pricing Basis A - Firm Requirement indicates that the initial requirement is for 12 Eye Tracking Glasses and 18 recording devices to assist in the event of equipment failure. Pricing Basis B are considered as options to purchase of additional items in any quantity up to the maximum quantity identified at a later date. Quantities in Pricing Basis A and Pricing Basis B should be considered as separate requirements.

Q4: Pricing Basis A and B – A3/B3 is requesting 18 additional batteries and wall chargers. Each Recording device will include 3 batteries and 1 wall charger, so A1/B1 will include 36 batteries and 12 wall chargers, and B2 will include either 18 batteries (if qty request is 6) or 54 batteries (if qty 18). Do you want this line item to stay in bid? A3 – is this a request for 18 additional wall chargers?

A4: The specifications for recording devices assume that recording devices include only 1 battery and no wall chargers (i.e., wall chargers are accounted for separately). That is, we are expecting a mandatory minimum of 18 recording devices, 18 batteries and 18 wall chargers (A2 + A3), and an optional additional 18 recording devices, 18 batteries and 12 wall chargers (B2 + B3 + B4), in any combination thereof, regardless of how they are bundled. If some of the requirements are bundled and account for another part of the requirement, that is fine, I would just ask the vendor to be explicit about that. If any particular item is bundled within another item, indicate a \$0.00 price. If not, provide pricing for that particular item.

Solicitation No. - N° de l'invitation
W7719-215523/A
Client Ref. No. - N° de réf. du client
6000515766

Amd. No. - N° de la modif.
002
File No. - N° du dossier
KIN-0-54170

Buyer ID - Id de l'acheteur
KIN930
CCC No./N° CCC - FMS No./N° VME

Q5: Pricing Basis B – B5 “Carrying Cases” Qty 2 – our company's hardware units include a carrying case for each hardware system – is this requirement for 2 additional empty cases?

A5: Yes, B5 indicates an option to purchase 2 additional cases.

Q6: A5/B6 – Controller software is included with purchase of each hardware unit and has unlimited activations so there is no itemized cost for this. B6 will only reflect the Analysis software cost.

A6: This is not a problem, please just indicate that this is the case.

Q7: Regarding Line item#4.2 RECORDING DEVICES FOR EYE TRACKING GLASSES “Weight (Maximum) Maximum: 500 grams”. Could you please clarify the weight of recorder with 650 grams will be acceptable.

A7: The mandatory maximum weight of 500g for the recording device was selected to reduce unnecessary burden on participants under the demanding conditions of field experiments. Additional weight may inhibit participants' movements or otherwise adversely impact their operational effectiveness in a field setting. Any weight for the recording device above 500g is not in compliance with mandatory requirements.