



RETURN BIDS TO:

RETOURNER LES SOUMISSIONS À:

Bid Receiving Public Works and Government
Services Canada/Réception des soumissions/Travaux
publics et Services gouvernementaux Canada
See herein for bid submission
instructions/

Voir la présente pour les
instructions sur la présentation
d'une soumission

NA

Manitoba

SOLICITATION AMENDMENT MODIFICATION DE L'INVITATION

The referenced document is hereby revised; unless otherwise
indicated, all other terms and conditions of the Solicitation
remain the same.

Ce document est par la présente révisé; sauf indication contraire,
les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

Vendor/Firm Name and Address

Raison sociale et adresse du
fournisseur/de l'entrepreneur

Issuing Office - Bureau de distribution

Public Works and Government Services Canada -
Western Region
Victory Building/Édifice Victory
Room 310/pièce 310
269 Main Street/269 rue Main
Winnipeg
Manitoba
R3C 1B3

Title - Sujet Solid State Fermentation System	
Solicitation No. - N° de l'invitation 31029-210706/C	Amendment No. - N° modif. 005
Client Reference No. - N° de référence du client 31029-210706	Date 2021-03-22
GETS Reference No. - N° de référence de SEAG PW-\$WPG-015-11167	
File No. - N° de dossier WPG-0-43093 (015)	CCC No./N° CCC - FMS No./N° VME
Solicitation Closes - L'invitation prend fin at - à 02:00 PM Central Daylight Saving Time CDT on - le 2021-04-06 Heure Avancée du Centre HAC	
F.O.B. - F.A.B. Plant-Usine: <input type="checkbox"/> Destination: <input checked="" type="checkbox"/> Other-Autre: <input type="checkbox"/>	
Address Enquiries to: - Adresser toutes questions à: Klassen, Chaz	Buyer Id - Id de l'acheteur wpg015
Telephone No. - N° de téléphone (204) 296-6920 ()	FAX No. - N° de FAX (418) 566-6167
Destination - of Goods, Services, and Construction: Destination - des biens, services et construction:	

Instructions: See Herein

Instructions: Voir aux présentes

Delivery Required - Livraison exigée	Delivery Offered - Livraison proposée
Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur	
Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur	
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date

This amendment 005 is being raised to address the following questions with regards to solicitation No 31029-210706/C

Part A: Questions and Answers

Q1. Item # 1 : The proposed system must offer a Trimode controller which can operate various fermentation vessels including Solid State, Jacketed, and Single vessels. **Do you mean single use vessels.?**

A1. This refers to Single Wall vessels, not Single use vessels.

Q2. Item # 3 : The proposed system must have the flexibility to accommodate different vessels, with volume range from 0.5 to 20 litres. The Solid State fermentation vessel must have a minimum volume of 5 litres and a maximum of 20 litres. **Benchtop system we are offering can be supplied with 1L, 2L, 5L or 10L. The Biostat B tower can support any two vessels, Biostat B-DCU can be expanded to support total 6 vessels and expansion can occur anytime. Please advise if that would be acceptable.**

A2. The proposal is for the procurement of a Solid State Fermentations system with the volume stated in Item #3. The system must have the ability for future expansion to include liquid cultures if required at a later date. The volumes proposed are acceptable, provided the mandatory requirement of Solid State Fermentation are met.

Q3. Question: Item # 7 : The proposed system must allow Solid State Fermentation vessel to be rotated from 0° - 90°, as well as 120° for harvesting. **In our proposed system there is an accessory to facilitate tipping vessel safely. Please advise if that would be acceptable.**

A3. The vessel must be able to be rotated on the stand that holds the vessel in the degrees stated. An added accessory that permits tipping is acceptable as long as the rotation requirements are also met.

Part B: Revisions

Page 1: Closing Date:

DELETE: March 23, 2021.

INSERT: April 6, 2021.

ALL OTHER TERMS AND CONDITIONS REMAIN THE SAME