



RETURN BIDS TO:

RETOURNER LES SOUMISSIONS À:

Bid Receiving Public Works and Government
Services Canada/Réception des soumissions Travaux
publics et Services gouvernementaux Canada
1713 Bedford Row
Halifax, N.S./Halifax, (N.É.)
Halifax
Nova Scotia
B3J 1T3
Bid Fax: (902) 496-5016

**SOLICITATION AMENDMENT
MODIFICATION DE L'INVITATION**

The referenced document is hereby revised; unless otherwise
indicated, all other terms and conditions of the Solicitation
remain the same.

Ce document est par la présente révisé; sauf indication contraire,
les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

Vendor/Firm Name and Address
Raison sociale et adresse du
fournisseur/de l'entrepreneur

Issuing Office - Bureau de distribution
Atlantic Region Acquisitions/Région de l'Atlantique
Acquisitions
1713 Bedford Row
Halifax, N.S./Halifax, (N.É.)
Halifax
Nova Scotia
B3J 1T3

Title - Sujet CFIA - Liquid Chromatograph ACIA - Chromatographe en phase liquide	
Solicitation No. - N° de l'invitation 39903-210743/A	Amendment No. - N° modif. 002
Client Reference No. - N° de référence du client 39903-21-0743	Date 2021-05-06
GETS Reference No. - N° de référence de SEAG PW-SHAL-104-6142	
File No. - N° de dossier HAL-0-84115 (104)	CCC No./N° CCC - FMS No./N° VME
Solicitation Closes - L'invitation prend fin at - à 02:00 PM Atlantic Daylight Saving Time ADT on - le 2021-05-20 Heure Avancée de l'Atlantique HAA	
F.O.B. - F.A.B. Plant-Usine: <input type="checkbox"/> Destination: <input checked="" type="checkbox"/> Other-Autre: <input type="checkbox"/>	
Address Enquiries to: - Adresser toutes questions à: Nowakowski, Leanne	Buyer Id - Id de l'acheteur hal104
Telephone No. - N° de téléphone (902) 403-7112 ()	FAX No. - N° de FAX (902) 496-5016
Destination - of Goods, Services, and Construction: Destination - des biens, services et construction:	

Instructions: See Herein

Instructions: Voir aux présentes

Delivery Required - Livraison exigée	Delivery Offered - Livraison proposée
Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur	
Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur	
Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie)	
Signature	Date

Solicitation Amendment 002 is being issued to provide additional questions and answers to date.

1) Questions and Answers:

5. I have a question in regards of 3. UHPLC Specifications: d. The UHPLC must contain a thermostatted column compartment that is capable of heating and cooling and provides a reproducible regulated temperature from 10 - 60 °C selectable in at least 1°C increments (25°C ambient temperature) Would : Capable of temperatures from 10 C below ambient temperature to 100 C be acceptable?

a. Yes "10° below ambient to 100°C" is acceptable

6. The specification indicates: 10 - 60 °C selectable in at least 1°C increments (25°C ambient temperature) – would 15°C (Ambient -10°C) to 100°C be acceptable?

a. Yes, "15°C (Ambient -10°C) to 100°C" is acceptable

7. In Amendment 001 you have indicated that 96 vials is not sufficient however would 162 vials be acceptable?

a. Yes, 162 vials is acceptable.

For further clarification, the 1.5-2 mL specification in 3.f refers to 12 x 32 mm vials.

8. Section 22 Warranty of the 2030 (2020-05-28), General Conditions. Historically, when selling our instrumentation to agencies through the PWGSC, we have found the automatic extension defined with subsection 7 to be egregious. Is the agency willing to revise Section 22 to RESERVE subsection 7? Such request has been acceptable in the past, but we are submitting during the Q/A period so not to jeopardize the entire submission upon evaluation.

a. At this time, PSPC will not revise subsection 7 of the warranty terms.

9. Our Legal Department reviewed the Higher Complexity terms read into this solicitation, and the products solicited do not fit within the usage criteria for that set of terms.. As stipulated within the SACC manual, we have always sold under the Medium Complexity Terms, since the solicited products and services are commercial off the shelf items. They are not any of the examples listed in the use for Higher Complexity standard terms. Could you please review if a correction should be made to reference the Medium Complexity (Goods) terms rather than the Higher Complexity?

a. Yes, PSPC is able to change the Terms and Conditions of the resulting Contract to Medium Complexity. Please see the itemized change below.

2) At 6.2.1 General Conditions:

Delete: In its entirety

Insert: 2010A (2020-05-28), General Conditions - Medium Complexity - Goods, apply to and form part of the Contract.

3) At 6.10 Priority of Documents:

Delete: (b) the general conditions 2030 (2020-05-28), General Conditions - Higher Complexity - Goods,

Insert: (b) the general conditions 2010A (2020-05-28), General Conditions - Medium Complexity-Goods,