



Amendment 1

Hyper-Spectral Sensing with Intense Ultrafast Laser Fields – Questions and Answers

Q1. I have a question regarding the tender for the tunable femtosecond laser in this tender. Specifically how, it should be “compatible” with the LightConversion Carbide CB3 pump laser? Is there a synchronization requirement?

A1. The solution must make use, in some form, of the existing LightConversion Carbide CB3 pump laser, which is equipped with an additional oscillator output. If the solution incorporates additional pump lasers, they must be synchronized in such a way that the outputs of the solution are also temporally synchronized.