



## RETURN BIDS TO:

## RETOURNER LES SOUMISSIONS À:

Bid Receiving Public Works and Government  
Services Canada/Réception des soumissions Travaux  
publics et Services gouvernementaux Canada  
1713 Bedford Row  
Halifax, N.S./Halifax, (N.É.)  
Halifax  
Nova Scotia  
B3J 1T3  
Bid Fax: (902) 496-5016

## SOLICITATION AMENDMENT MODIFICATION DE L'INVITATION

The referenced document is hereby revised; unless otherwise  
indicated, all other terms and conditions of the Solicitation  
remain the same.

Ce document est par la présente révisé; sauf indication contraire,  
les modalités de l'invitation demeurent les mêmes.

### Comments - Commentaires

Vendor/Firm Name and Address  
Raison sociale et adresse du  
fournisseur/de l'entrepreneur

Issuing Office - Bureau de distribution  
Atlantic Region Acquisitions/Région de l'Atlantique  
Acquisitions  
1713 Bedford Row  
Halifax, N.S./Halifax, (N.É.)  
Halifax  
Nova Scot  
B3J 1T3

<b>Title - Sujet</b> Signal Generator DND (FMFCS) has a requirement for 2 Signal Generators, and related equipment.	
<b>Solicitation No. - N° de l'invitation</b> W355B-215215/A	<b>Amendment No. - N° modif.</b> 006
<b>Client Reference No. - N° de référence du client</b> W355B-21-5215	<b>Date</b> 2021-10-12
<b>GETS Reference No. - N° de référence de SEAG</b> PW-SHAL-208-11336	
<b>File No. - N° de dossier</b> HAL-1-87029 (208)	<b>CCC No./N° CCC - FMS No./N° VME</b>
<b>Solicitation Closes - L'invitation prend fin</b> <b>at - à 02:00 PM</b> Atlantic Daylight Saving Time ADT <b>on - le 2021-10-21</b> Heure Avancée de l'Atlantique HAA	
<b>F.O.B. - F.A.B.</b> <b>Plant-Usine:</b> <input type="checkbox"/> <b>Destination:</b> <input checked="" type="checkbox"/> <b>Other-Autre:</b> <input type="checkbox"/>	
<b>Address Enquiries to: - Adresser toutes questions à:</b> LeBlanc, JoAnne	<b>Buyer Id - Id de l'acheteur</b> hal208
<b>Telephone No. - N° de téléphone</b> (902) 225-6730 ( )	<b>FAX No. - N° de FAX</b> (902) 496-5016
<b>Destination - of Goods, Services, and Construction:</b> <b>Destination - des biens, services et construction:</b>	

Instructions: See Herein

Instructions: Voir aux présentes

<b>Delivery Required - Livraison exigée</b>	<b>Delivery Offered - Livraison proposée</b>
<b>Vendor/Firm Name and Address</b> <b>Raison sociale et adresse du fournisseur/de l'entrepreneur</b>	
<b>Telephone No. - N° de téléphone</b> <b>Facsimile No. - N° de télécopieur</b>	
<b>Name and title of person authorized to sign on behalf of Vendor/Firm</b> <b>(type or print)</b> <b>Nom et titre de la personne autorisée à signer au nom du fournisseur/</b> <b>de l'entrepreneur (taper ou écrire en caractères d'imprimerie)</b>	
<b>Signature</b>	<b>Date</b>

---

**THIS AMENDMENT NO. 006 TO THE REQUEST FOR PROPOSAL IS BEING RAISED  
TO ADDRESS THE BIDDER'S QUESTIONS.**

**REFER TO:**

**Under Annex A – Statement of Requirement**

**Mandatory Specifications section:**

**Question 1:**

**Under Section 2.1. Microwave Analog Signal Generators, Qty. (2)**

- **Statement:** 2.1.10. **Must** have a front cover
  - **Question:** Would you accept a carrying case for better overall protection of the instrument ?

**Answer:** **YES**

**Question 2:**

**Under Section 2.2 115 dB Step Attenuators, - Qty. (2)**

- **Statement:** 2.2.1. **Must** have Attenuator hold range -15 dBm to maximum specified output power with step attenuator in 0 dB state; can be offset using mechanical attenuator
  - **Question:** Would you accept an electronic attenuator with the same performance in lieu of a mechanical attenuator ?

An electronic attenuator ensures wear and tear free switching which can prolong internal components life span, and provides faster level settling times.

**Answer:** **YES**

**Question 3:**

**Under Section 2.3 High output power, - Qty. (2)**

- **Statement:** 2.3.1. **Must** provide maximum output power of 9kHz to 3.2GHz +23dBm, >3.2 to 13 GHz +20dBm, 13 to 20 GHz +19 dBm
  - **Question:** a) Do you have a fixed requirement for power levels up to +23dBm below 1MHz ?

**Answer:** **NO**

- **Question:** b) Would the following specifications meeting your requirements:
  - 8 KHz to ≤ 20 KHz +8 dBm
  - 20 KHz to ≤ 1 MHz +13 dBm
  - 1 MHz to ≤ 8 MHz +25 dBm
  - 8 MHz to ≤ 6 GHz +28 dBm
  - 6 GHz to ≤ 18 GHz +27 dBm
  - 18 GHz to 20 GHz +24 dBm

**Answer:** **YES**

***ALL OTHER TERMS AND CONDITIONS REMAIN UNCHANGED***