



**RETURN BIDS TO:**

**RETOURNER LES SOUMISSIONS À:**

Bid Receiving Public Works and Government  
Services Canada/Réception des soumissions Travaux  
publics et Services gouvernementaux Canada  
See herein for bid submission  
instructions/

Voir la présente pour les  
instructions sur la présentation  
d'une soumission

NA  
NA

**SOLICITATION AMENDMENT  
MODIFICATION DE L'INVITATION**

The referenced document is hereby revised; unless otherwise  
indicated, all other terms and conditions of the Solicitation  
remain the same.

Ce document est par la présente révisé; sauf indication contraire,  
les modalités de l'invitation demeurent les mêmes.

**Comments - Commentaires**

**Vendor/Firm Name and Address**  
**Raison sociale et adresse du**  
**fournisseur/de l'entrepreneur**

**Issuing Office - Bureau de distribution**  
Public Works and Government Services Canada - Pacific  
Region  
219 - 800 Burrard Street  
800, rue Burrard, pièce 219  
Vancouver, BC V6Z 0B9

<b>Title - Sujet</b> Battery Test System	
<b>Solicitation No. - N° de l'invitation</b> 31019-218448/A	<b>Amendment No. - N° modif.</b> 001
<b>Client Reference No. - N° de référence du client</b> 31019-218448	<b>Date</b> 2021-11-17
<b>GETS Reference No. - N° de référence de SEAG</b> PW-\$VAN-524-9072	
<b>File No. - N° de dossier</b> VAN-1-44120 (524)	<b>CCC No./N° CCC - FMS No./N° VME</b>
<b>Solicitation Closes - L'invitation prend fin</b> <b>at - à 02:00 PM</b> Pacific Standard Time PST <b>on - le 2021-11-29</b> Heure Normale du Pacifique HNP	
<b>F.O.B. - F.A.B.</b> <b>Plant-Usine:</b> <input type="checkbox"/> <b>Destination:</b> <input checked="" type="checkbox"/> <b>Other-Autre:</b> <input type="checkbox"/>	
<b>Address Enquiries to: - Adresser toutes questions à:</b> Yamamoto, Albert	<b>Buyer Id - Id de l'acheteur</b> van524
<b>Telephone No. - N° de téléphone</b> (604) 562-8773 ( )	<b>FAX No. - N° de FAX</b> ( ) -
<b>Destination - of Goods, Services, and Construction:</b> <b>Destination - des biens, services et construction:</b>	

**Instructions: See Herein**

**Instructions: Voir aux présentes**

<b>Delivery Required - Livraison exigée</b>	<b>Delivery Offered - Livraison proposée</b>
<b>Vendor/Firm Name and Address</b> <b>Raison sociale et adresse du fournisseur/de l'entrepreneur</b>	
<b>Telephone No. - N° de téléphone</b> <b>Facsimile No. - N° de télécopieur</b>	
<b>Name and title of person authorized to sign on behalf of Vendor/Firm</b> <b>(type or print)</b> <b>Nom et titre de la personne autorisée à signer au nom du fournisseur/</b> <b>de l'entrepreneur (taper ou écrire en caractères d'imprimerie)</b>	
<b>Signature</b>	<b>Date</b>

## Questions (Q) and Answers (A):

Q1: Are you looking for a single box application to meet all your measurement requirements?

A1: See A9, dimensions. We are not interested in placing 16 boxes around the lab. Our vision was something like a rack mounted system. As an example it could be:

- A couple of 8-channel multi-stats in a rack
- It could be 16 off-the-shelf potentiostat/FRA units in a rack
- It could be a proprietary 16 channel system

Products supplied by Maccor, Arbin, Gamry, Biologic, Solartron and other similar international suppliers for examples of product offerings.

Q2: To us, it looks like you will require multiple units to complete your measurements and readings – if that is the case or acceptable; we need to discuss this in further detail.

A2: We require a single system, dimensions in A9, with 16 channels described in B2 through B4.

Q3: Is there a specific Manufacturer that you are working with directly on this? If so, can you please confirm the name of the Manufacturer

A3: No.

Q4: The NPP states “All Mandatory specifications must be satisfied to be successfully awarded this opportunity”. Do each of the mandatory specifications truly need to be satisfied as specified or is there any additional flexibility?

A4: No flexibility; meet or exceed. Refer to the Request for Proposal (RFP) and not just the NPP. All mandatory criteria in the RFP must be satisfied.

Q5: Does the EIS capability need to be ran simultaneously across each of the 8 channels or do these 8 channels just need to be able to use EIS?

Solicitation No. - N° de l'invitation  
31019-218448/A  
Client Ref. No. - N° de réf. du client  
31019-218448

Amd. No. - N° de la modif.  
001  
File No. - N° du dossier

Buyer ID - Id de l'acheteur  
VAN524  
CCC No./N° CCC - FMS No./N° VME

---

A5: Specifications states simultaneous

Q6: Does the lowest current range need to reach 500 uA (within 20%) or is there any flexibility for this?

A6: Specifications as stated

Q7: Can the data be stored in a local SQL, vs on the computer in plain text format, and then be exported to either CSV or excel files?

A7: No – we need to be able to read the data in real time for safety reasons.