



**RETURN BIDS TO:**

**RETOURNER LES SOUMISSIONS À:**

PWGSC/TPSGC Acquisitions Bid Receiving

Box/Boîte de Réception des Soumissions

Bid Receiving Box/Boîte de Récepti

1st Floor/1<sup>ère</sup> étage, Suite 1212

100-1045 Main Street

Moncton

New Brunswick

E1C 1H1

Bid Fax: (506) 851-6759

**SOLICITATION AMENDMENT  
MODIFICATION DE L'INVITATION**

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Solicitation remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de l'invitation demeurent les mêmes.

**Comments - Commentaires**

**Vendor/Firm Name and Address**

Raison sociale et adresse du  
fournisseur/de l'entrepreneur

**Issuing Office - Bureau de distribution**

Acquisitions NB/PEI (Moncton Office) – Bureau

d'acquisitions N.-B./Î.-P.-É. (Moncton)

1045 Main Street / 1045, rue Main

Moncton

New Bruns

E1C 1H1

<b>Title - Sujet</b> Biological Safety Cabinet Biological Safety Cabinet(BSC) with base stand Class II type A2&Class II type B2	
<b>Solicitation No. - N° de l'invitation</b> 39903-220633/A	<b>Amendment No. - N° modif.</b> 001
<b>Client Reference No. - N° de référence du client</b> 39903-220633	<b>Date</b> 2021-11-30
<b>GETS Reference No. - N° de référence de SEAG</b> PW-\$MCT-033-6076	
<b>File No. - N° de dossier</b> MCT-1-44077 (033)	<b>CCC No./N° CCC - FMS No./N° VME</b>
<b>Solicitation Closes - L'invitation prend fin</b> <b>at - à 02:00 PM</b> Atlantic Standard Time AST <b>on - le 2021-12-09</b> Heure Normale de l'Atlantique HNA	
<b>F.O.B. - F.A.B.</b> <b>Plant-Usine:</b> <input type="checkbox"/> <b>Destination:</b> <input checked="" type="checkbox"/> <b>Other-Autre:</b> <input type="checkbox"/>	
<b>Address Enquiries to: - Adresser toutes questions à:</b> Young (MCT), Leesa	<b>Buyer Id - Id de l'acheteur</b> mct033
<b>Telephone No. - N° de téléphone</b> (506) 871-1716 ( )	<b>FAX No. - N° de FAX</b> (506) 851-6759
<b>Destination - of Goods, Services, and Construction:</b> <b>Destination - des biens, services et construction:</b>	

Instructions: See Herein

Instructions: Voir aux présentes

<b>Delivery Required - Livraison exigée</b>	<b>Delivery Offered - Livraison proposée</b>
<b>Vendor/Firm Name and Address</b> <b>Raison sociale et adresse du fournisseur/de l'entrepreneur</b>	
<b>Telephone No. - N° de téléphone</b> <b>Facsimile No. - N° de télécopieur</b>	
<b>Name and title of person authorized to sign on behalf of Vendor/Firm</b> <b>(type or print)</b> <b>Nom et titre de la personne autorisée à signer au nom du fournisseur/</b> <b>de l'entrepreneur (taper ou écrire en caractères d'imprimerie)</b>	
<b>Signature</b>	<b>Date</b>

---

## Solicitation Amendment No.001

**Title:** Biological Safety Cabinet (BSC) with base stand Class II type A2 and Class II type B2

*This solicitation is hereby amended to provide the following questions and answers:*

**Question # 1:** The Class IIB2, is that be connected to a existing ducting in place or are they requiring installation of ducting and blower on roof normally handled separately by a mechanical contractor? The Class II Type B2 BSC will be installed with a hard-ducted canopy connection and will exhaust air to the outside environment.

**Answer #1:** The Class II B2 BSC will be connected to an existing hard-ducted canopy connection.

**Question # 2:** Annex A, 4.0 Requirements, Item#2.1: 3. Class 11 Type B2 hood sash height is listed as 10" height but the industry standard for Type 11 B2 is 8" sash height due to air flow demands of a 10" sash opening would require. Will client accept 8" sash opening?

**Answer #2:** Yes, we will accept a 8" opening

**Question # 3:** Annex A, 4.0 Requirements, Installation: Need clarification oof installation of the B2 Hood. We spoke to a ConTest Air who is a service supplier for BSC hoods and they will do everything but they do not make the connection to the duct which would be responsibility of the building maintenance. They will lift the hood up in place and then Bldg would make the connection. ConTest would then complete certification? Is this acceptable?

**Answer #3:** Yes, Charlottetown Lab building maintenance will make the connection.

**Question # 4:** For section 1.5, will you accept a unit that does not provide the life of the filter information and only provides alarms for blower power- this prompts you to change the filter which should only happen once in the lifetime of the equipment (probably)?

**Answer # 4:** Yes

**Question # 5:** For section 1.7, will you accept aluminum covers?

**Answer # 5:** No

**Question # 6:** For section 1.9, will you accept a different design armrest that has multiple positions for forearm rest?

**Answer # 6:** Yes

Solicitation No. - N° de l'invitation  
39903-220633/A  
Client Ref. No. - N° de réf. du client  
39903-220633

Amd. No. - N° de la modif.  
001  
File No. - N° du dossier

Buyer ID - Id de l'acheteur  
MCT033  
CCC No./N° CCC - FMS No./N° VME

---

If your bid has already been forwarded and you wish to revise same, this revision should be sent either in a sealed envelope and mailed to the above address or by facsimile (506) 851-6759 and reach the undersigned before the appropriate closing date. The solicitation number and the closing date are to be shown on the outside of the sealed envelope or on the facsimile transmission.

All other terms and conditions of the solicitation document remain unchanged remain unchanged.

All enquiries concerning this amendment are to be forwarded to:

Name: Leesa Young  
Telephone No: (506) 871-1716  
Facsimile No: (506-851-6759  
Email: leesa.young@pwgsc-tpsgc.gc.ca