



RETURN BIDS TO:

RETOURNER LES SOUMISSIONS À:

Bid Receiving - PWGSC / Réception des soumissions -
TPSGC

See Herein / Voir ci-joint

Bid Fax: (819) 997-9776

**SOLICITATION AMENDMENT
MODIFICATION DE L'INVITATION**

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Solicitation remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de l'invitation demeurent les mêmes.

Comments - Commentaires

Vendor/Firm Name and Address

Raison sociale et adresse du
fournisseur/de l'entrepreneur

Issuing Office - Bureau de distribution

Scientific, Medical and Photographic Division /
Division de l'équipement scientifique, des produits
photographiques et pharmaceutiques
L'Esplanade Laurier
140 O'Connor Street,
East Tower, 7th Floor
Ottawa
Ontario
K1A 0S5

| | |
|--|---|
| Title - Sujet Load Cells Load Cells | |
| Solicitation No. - N° de l'invitation T8009-210227/A | Amendment No. - N° modif. 001 |
| Client Reference No. - N° de référence du client T8009-210227 | Date 2022-06-01 |
| GETS Reference No. - N° de référence de SEAG PW-\$\$PV-867-81200 | |
| File No. - N° de dossier pv867.T8009-210227 | CCC No./N° CCC - FMS No./N° VME |
| Solicitation Closes - L'invitation prend fin at - à 02:00 PM Eastern Daylight Saving Time EDT on - le 2022-06-09 Heure Avancée de l'Est HAE | |
| F.O.B. - F.A.B. Plant-Usine: <input type="checkbox"/> Destination: <input type="checkbox"/> Other-Autre: <input type="checkbox"/> | |
| Address Enquiries to: - Adresser toutes questions à: Golab, Lisa | Buyer Id - Id de l'acheteur pv867 |
| Telephone No. - N° de téléphone (343) 574-2635 () | FAX No. - N° de FAX () - |
| Destination - of Goods, Services, and Construction: Destination - des biens, services et construction: | |

Instructions: See Herein

Instructions: Voir aux présentes

| | |
|--|--|
| Delivery Required - Livraison exigée | Delivery Offered - Livraison proposée |
| Vendor/Firm Name and Address Raison sociale et adresse du fournisseur/de l'entrepreneur | |
| Telephone No. - N° de téléphone Facsimile No. - N° de télécopieur | |
| Name and title of person authorized to sign on behalf of Vendor/Firm (type or print) Nom et titre de la personne autorisée à signer au nom du fournisseur/ de l'entrepreneur (taper ou écrire en caractères d'imprimerie) | |
| Signature | Date |

Amendment 001 has been raised to provide answers to questions from industry and modify the bid solicitation.

A. QUESTIONS AND ANSWERS

- Q1.** At Annex A:
"Please advise the type of load cell is required."
- A1.** The load cell must be a 6-component force-torque load cell. Please see Modification to the RFP, herein.
- Q2.** At Annex A:
"Will the load cell be used in tension, compression, both or shear?"
- A2.** Transport Canada will use the load cells in tension and compression. Please see Modification to the RFP, herein.
- Q3.** At Annex A:
"Is the load cell a compression load cell, a through hole load cell, etc."
- A3.** The load cell must have one mounting surface on both ends to allow for installation in a balance structure. Please see Modification to the RFP, herein.

B. MODIFICATIONS

1. At Annex A, Part 1, Section 2.0 - MANDATORY TECHNICAL REQUIREMENT:

Insert the following Mandatory Technical Requirements:

- 2.6 The load cells must be a 6-component force-torque load cell and have tension and compression.
- 2.7 The load cells must have one mounting surface on both ends to allow for installation in a balance structure.

2. At Annex A, Part 2.1 - MANDATORY TECHNICAL EVALUATION CRITERIA:

Insert the following Mandatory Technical Evaluation Criteria:

| ITEM | CRITERIA | REFERENCE TO SUBSTANTIATION IN THE TECHNICAL BID. |
|-----------|--|---|
| M6 | The load cells must be a 6-component force-torque load cell to be used in tension and compression. The bidder must demonstrate compliance by providing document such as literature, specifications, catalogues as supporting documents. | |
| M7 | The load cells must have one mounting surface on both ends to allow for installation in a balance structure. | |

Solicitation No. - N° de l'invitation
T8009-210227/A
Client Ref. No. - N° de réf. du client
T8009-210227

Amd. No. - N° de la modif.
001
File No. - N° du dossier
pv867.T8009-210227

Buyer ID - Id de l'acheteur
pv867
CCC No./N° CCC - FMS No./N° VME

| | | |
|--|--|--|
| | The bidder must demonstrate compliance by providing document such as literature, specifications, catalogues as supporting documents. | |
|--|--|--|

3. The closing date of the bid solicitation has been extended as follows:

DELETE: 2022-06-02 14:00 Eastern Daylight Time (EDT)

INSERT: 2022-06-09 14:00 Eastern Daylight Time (EDT)

ALL OTHER TERMS AND CONDITIONS REMAIN UNCHANGED