

Trade Agreement: Canadian Free Trade Agreement (CFTA)

Tendering Procedures: All pre-qualified suppliers may bid.

Competitive Procurement Strategy: Lowest/Lower Bids

Comprehensive Land Claim Agreement: No

Nature of Requirements:

Dairy Fridge
2023-00028C
Alexander Lisowski
613-285-6119
Alexander.lisowski@inspection.gc.ca

THIS REQUEST IS RESERVED FOR HOLDERS OF SUPPLY ARRANGEMENT # E60PV-19EQUI ONLY, FOR THE PURCHASE OF LABORATORY AND SCIENTIFIC EQUIPMENT, PARTS AND ACCESSORIES, SERVICES AND SUPPLIES.

Only Suppliers currently pre-qualified on Supply Arrangement E60PV-19EQUI have been invited to bid.

As a requirement of the Supply Arrangement, this notice is published on Buy and Sell Canada for a period of 40 calendar days. The closing date published on this notice identifies how long the notice will be published. For the closing date of any solicitation under the supply arrangement, invited suppliers should refer to the solicitation documents.

Suppliers that do not have a Supply Arrangement for the supply of Laboratory and Scientific Equipment, Parts and Accessories, Services and Supplies with Public Works and Government Services Canada, cannot submit a bid. Any bids received from suppliers not pre-qualified on the Supply Arrangement will not be evaluated.

Suppliers may qualify under Supply Arrangement E60PV-19EQUI for Laboratory and Scientific Equipment, Parts and Accessories, Services and Supplies, at any time. Interested suppliers should download solicitation document E60PV-19EQUI /C from Buyandsell.gc.ca and submit a response as per the requirements of the Request for Supply Arrangement.

Description:

The CFIA Sidney Laboratory has a requirement for a large dairy fridge and options to purchase 3 additional units over the next 3 years.

The mandatory technical criteria are:

CRITERIA	
M1	Dimensions: Exterior Dimensions: Must be a Maximum of 81"W x 34"D x 79"H (without casters)Note:Max height of 79" is without castors on. Capacity: Fridge must have minimum capacity of 65cu.ft Weight: Fridge must be a maximum 760 lbs
M2	Exterior: Must be made of a durable non-peel or chip, white laminated vinyl or powder-coated steel Interior: Must be made of Stainless steel or powder-coated steel
M3	Doors: Material: Must be made of glass

	<p>Pane: Glass must be made of double-pane glass</p> <p>Function: Doors must open by swinging or sliding</p> <p>Energy Efficiency: Glass must be energy efficient</p> <p>Seal: Door seal must be a positive seal</p>
M4	Temperature: Fridge must be able to maintain 4 degrees C
M5	Display: Fridge must have an Exterior digital temperature display with alarm function
M6	Refrigerant: Fridge must use R290 hydrocarbon (no ozone depletion potential) or other natural refrigerant that is environmentally friendly.
M7	Shelving: The Fridge must have a minimum of 9 (to a maximum of 12) adjustable heavy-duty shelves
M8	<p>Electrical Aspects:</p> <p>Compressor: Compressor must be a minimum of ½ hp</p> <p>Electrical: Electrical components must be 115 V</p>

The deliverables must include delivery, installation and warranty as outlined in the bid solicitation.

The responsive bidder already holding a Supply arrangement for Laboratory and Scientific Equipment, Parts and Accessories, Services and Supplies from Public Works and Government Services Canada who meets the mandatory technical criteria with the lowest aggregate price will be recommended for contract award.