



**REQUEST FOR PROPOSAL /
DEMANDE DE PROPOSITION**

**RETURN BIDS TO /
RETOURNER LES SOUMISSIONS À:**

Bid Receiving – PWGSC / Réception des
soumissions - TPSGC
11 Laurier St. / 11 rue Laurier
Place du Portage, Phase III
Core 0B2 / Noyau 0B2
Gatineau
Québec
K1A 0S5

Proposal To: National Defence Canada

We hereby offer to sell to Her Majesty the Queen in
right of Canada, in accordance with the terms and
conditions set out herein, referred to herein or
attached hereto, the goods and services listed herein
and on any attached sheets at the price(s) set out
therefore.

Proposition à : Défense nationale Canada

Nous offrons par la présente de vendre à Sa Majesté la
Reine du chef du Canada, aux conditions énoncées ou
incluses par référence dans la présente et aux annexes
ci-jointes, les biens et services énumérés ici et sur
toute feuille ci-annexée, au(x) prix indique(s).

**Solicitation Closes /
L'invitation prend fin:**

At / à : 14:00 EST

On / le : February 7th 2023

Title / Titre: Vibration Table Replacement/ REEMPLACEMENT DE LA TABLE VIBRANTE		Solicitation No / No de l'invitation: W8472-215850/A AMD 006	
Date of Solicitation / Date de l'invitation: 2022-10-21			
Address Enquiries to – Adresser toutes questions à: Dalton Sicard 101 Colonel by Drive Ottawa ON D Mar P 2 K1A 0K2 dalton.sicard@forces.gc.ca			
Telephone No. / N° de téléphone:		FAX No / No de fax: N/A	
Destination: 9401 Wanklyn St., Lasalle QC H8R 1Z2 Canada			

Instructions:

Municipal taxes are not applicable. Unless otherwise specified herein all prices quoted must include all applicable Canadian customs duties, GST/HST, excise taxes and are to be delivered Delivery Duty Paid including all delivery charges to destination(s) as indicated. The amount of the Goods and Services Tax/Harmonized Sales Tax is to be shown as a separate item.

Instructions:

Les taxes municipales ne s'appliquent pas. Sauf indication contraire, les prix indiqués doivent comprendre les droits de douane canadiens, la TPS/TVH et la taxe d'accise. Les biens doivent être livrés «rendu droits acquittés», tous frais de livraison compris, à la ou aux destinations indiquées. Le montant de la taxe sur les produits et services/taxe de vente

Delivery required / Livraison exigée:	Delivery offered / Livraison proposée:
Vendor Name and Address / Raison sociale et adresse du fournisseur:	
Name and title of person authorized to sign on behalf of vendor (type or print) / Nom et titre de la personne autorisée à signer au nom du fournisseur (caractère d'imprimerie):	
Name / Nom: _____	Title / Titre: _____
Signature: _____	Date: _____

Solicitation No. - N° de l'invitation
W8472-215850/A
Client Ref. No. - N° de réf. du client
W8472-215850/A

Amd. No. - N° de la modif.
006
File No. - N° du dossier
W8472-215850

Buyer ID - Id de l'acheteur
Q08
CCC No./N° CCC - FMS No./N° VME

Request for Proposal (RFP) W8472-215850/A - Vibration Table Replacement

Solicitation amendment 006 is proposed to extend the solicitation closing date and answer bidder's questions.

DELETE

Solicitation Closes – L'invitation prend fin
At – à : 14:00 EST
On - le : January 31st 2023

INSERT

Solicitation Closes – L'invitation prend fin
At – à : 14:00 EST
On - le : Febuary 7th 2023

Question 1:

In the last Amendment 005 attached, you answered the below question regarding the reaction mass:

27. Would you like a separate line item in the quote for engineering hours devoted to the design assistance of the reaction mass?

Yes, but it will not be a part of the evaluation process under this bid.

Can you please answer the same question as it relates to ALL aspects of Project Management, in other words, would like to see this as a separate line item on the quote and will this be (or not be) part of the evaluation process under this bid? Perhaps the answer is buried somewhere in the tender details but it would be best to get the best clarification possible on this.

Answer 1:

If the reaction mass is not an integral part of the shaker design (i.e. the bidder is not delivering both), then, as per the RFP Annex A para 10.4.1.2: NETE will coordinate, if not integral to the Shaker design, the design and construction of a new sub-foundation reaction mass (seismic mass) for isolation with the building structure and any environmental noise.

So for that matter, we would a separate line item in the quote for engineering hours devoted to the design assistance of the reaction mass. And that is not a part of the evaluation process.

The rest of the Project Management (as outlined in the RFP Annex F Table 2) will be subject to the evaluation.