

Amendment 1

Quantum Imaging – Questions and Answers

- Q1.** Would it be acceptable if the time for the system to complete a full 1 degree x 0.2 degree image was on the order of 10min?
- A1.** Yes, that is acceptable. We are aware this is a challenge and we are aiming at the shortest acquisition time, without imposing a quantitative limit.

Modification n° 001

Image Quantique – Questions et réponses

- Q1.** Serait-il acceptable que le temps nécessaire au système pour compléter une image complète de 1 degré x 0,2 degré soit de l'ordre de 10 minutes?
- R1.** Oui, c'est acceptable. Nous sommes conscients qu'il s'agit d'un défi et nous visons le délai d'acquisition le plus court, sans imposer de limite quantitative.