

Amendment No. 004

This amendment is being raised to 1) provide a response to Request for Clarification.

1) Response to Request for Clarification

Question 1.

Q1. In Annex A - STATEMENT OF REQUIREMENT: article 1. Background: for station B, it is mentioned: "Upgrade Humidification with Rapid Response and Consumables Kit." For Upgrade Humidification with Rapid Response, we offer a dynamic humidity control for the anode and cathode that can increase and decrease the dew point with a fast response (30% - 80% dew point will reach within 10 sec under current temperature.) is it acceptable?

R1. Yes, it's acceptable.

All other terms and conditions of the solicitation remain the same.