

Questions and Answers / Questions et Réponses

No./N°

1

Project Description / Description de projet		
Acquisition of Constant Climate Chamber		
Acquisition de chambre à climat constant		
Solicitation No./ N° de sollicitation	Departmental Representative / Représentant Ministériel	Date
24-58045-B	Stéphane Lajoie, stephane.lajoie@nrc-cnrc.gc.ca	Jun 26, 2024 26 juin 2024
Notice:		Avis:
This Q&A shall form part of the tender documents and all conditions shall apply and be read in conjunction with the RFP.		Cet Q&R fait partie intégrale des dossiers d'appel d'offres; toutes les conditions énoncées doivent être lues et appliquées en conjonction avec la DDP.

*Le français suit

Q1. For this Inner size 1040x1200x550mm, external size 1300x1800x820mm is small and not workable. Will the external size 1600x2100x1350mm be acceptable?

A1. Those external dimensions are not acceptable, as it exceeds the width of the elevator door we need to move it through.

=====

Q1. Pour cette taille intérieure de 1040x1200x550mm, la taille extérieure de 1300x1800x820mm est petite et ne peut pas être utilisée. La taille extérieure 1600x2100x1350mm sera-t-elle acceptable ?

R1. Ces dimensions extérieures ne sont pas acceptables, car elles dépassent la largeur de la porte de l'ascenseur par laquelle nous devons le faire passer.