

### **Amendment No. 001**

This amendment is being raised to provide a response to Request for Clarification.

#### Question 1.

Q1. Would you please confirm if the system must be CSA certified with label?

R1. Yes. All equipment is to be supplied to NRC as per the terms and conditions:  
B1501C (2018-06-21) Electrical equipment.

#### Question 2.

Q2. Can you please clarify with the researcher the specifications are for a Fuel Cell test station and not electrolysis?

R2. The specifications are for a CO<sub>2</sub> electrolysis test station. Refer to Page 18, Section 5: Fluid Control, which is related to the electrolyte control in an electrolysis test station.

**All other terms and conditions of the solicitation remain the same.**