



National Defence

Défense nationale

National Defence Headquarters
Ottawa, Ontario
K1A 0K2

Quartier général de la Défense nationale
Ottawa (Ontario)
K1A 0K2

SOLICITATION AMENDMENT / MODIFICATION DE L'INVITATION

RETURN BIDS TO / RETOURNER LES SOUMISSIONS À:

Attn : Thomas Kardaras
Department of National Defence
Director Services Contracting (D Svcs C 3)
101 Colonel By Drive
Ottawa, Ontario K1A 0K2
Email: Thomas.Kardaras@forces.gc.ca

The referenced document is hereby revised; unless otherwise indicated, all other terms and conditions of the Solicitation remain the same.

Ce document est par la présente révisé; sauf indication contraire, les modalités de l'invitation demeurent les mêmes

Title / Titre Phlebotomy Chairs	
Solicitation No. / N° de l'invitation W6369-23-A071A	Amendment No. / N° de la modification 001
Date of Amendment / Date de la modification 20-September-2023	
Address Enquiries to / Adresser toutes questions à: Attn : Thomas Kardaras Department of National Defence Director Services Contracting (D Svcs C 3) 101 Colonel By Drive Ottawa, Ontario K1A 0K2 Email: Thomas.Kardaras@forces.gc.ca	
Telephone No. / N° de téléphone	FAX No. / N° de fax
Destination Central Medical Equipment Depot (CMED) 105 Montgomery Road, Building BB104A Canadian Forces Base Petawawa Petawawa, ON K8K 2X3	

Instructions: Municipal taxes are not applicable. Unless otherwise specified herein all prices quoted must include all applicable Canadian customs duties, GST/HST, excise taxes and are to be delivered Delivery Duty Paid including all delivery charges to destination(s) as indicated. The amount of the Goods and Services Tax/Harmonized Sales Tax is to be shown as a separate item.

Instructions : Les taxes municipales ne s'appliquent pas. Sauf indication contraire, les prix indiqués doivent comprendre les droits de douane canadiens, la TPS/TVH et la taxe d'accise. Les biens doivent être livrés « rendu droits acquittés », tous frais de livraison compris, à la ou aux destinations indiquées. Le montant de la taxe sur les produits et services/taxe de vente harmonisée doit être indiqué séparément.

Solicitation Closes / L'invitation prend fin: At / à: 02:00 PM Eastern Daylight Time (EDT) On / le: 29-September-2023
--

Delivery Required / Livraison exigée	Delivery Offered / Livraison proposée
Vendor Name and Address / Raison sociale et adresse du fournisseur 	
Name and title of person authorized to sign on behalf of vendor (type or print) / Nom et titre de la personne autorisée à signer au nom du fournisseur (caractère d'imprimerie) Name – Nom _____ Title – Titre _____ Signature _____ Date _____	



National Defence

Défense nationale

National Defence Headquarters
Ottawa, Ontario
K1A 0K2

Quartier général de la Défense nationale
Ottawa (Ontario)
K1A 0K2

AMENDMENT 001 TO SOLICITATION NUMBER W6369-23-A071A IS RAISED TO:

1. Answer bidder questions; and
2. Modify ANNEX A – REQUIREMENT, Section 6.2 of SPECIFICATIONS.

QUESTIONS AND ANSWERS:

Number	Questions/Answers/Amendments
Question 1:	<p>Due to health risks, the fire-retardant process used for CAL 133 is no longer allowed in California. CAL 117 is standard for upholstered furniture flammability in the US and Canada.</p> <p>Would you please replace item 6.2.2, “Meet CA133 Fire Code Specifications for flame retardancy,” with Meet CAL117 Fire Code Specifications for flame retardancy”?</p>
Answer 1:	<p>Sub-section 6.2 will be modified to include CAL117 as per this amendment 001.</p>

SOLICITATION NUMBER W6369-23-A071A IS HEREBY AMENDED AS FOLLOWS:

1. **DELETE** ANNEX A – REQUIREMENT Section 6.2 in its entirety, and **REPLACE** with the following:
 - 6.2 The reclining phlebotomy chair’s upholstery must:
 - 6.2.1 Have a coating that is antimicrobial and stain resistant. For example, this could be a coating using silver ion antimicrobial technology; and
 - 6.2.2 Meet CAL133 or CAL177 Fire Code Specifications for flame retardancy.

ALL OTHER TERMS AND CONDITIONS REMAIN THE SAME